



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116269

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	09/13/2021	10:45
2	M3706-08		1	10	100.00	>212.0	>212.0	09/13/2021	11:10
3	M3706-08DUP		1	10	100.00	>212.0	>212.0	09/13/2021	11:35
4	M3706-10		1	12	100.00	>212.0	>212.0	09/13/2021	12:05
5	M3706-12		1	10	100.00	>212.0	>212.0	09/13/2021	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp91321 WorkList ID : 152184 Department : Wet-Chemistry Date : 09-13-2021 09:04:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/17/2021	Water	M3706-08	C Flash Point	Cool 4 deg C	PSEG03	H41	33442	09/10/2021	1010B
09/17/2021	Water	M3706-10	A Flash Point	Cool 4 deg C	PSEG03	H41	36-40	09/10/2021	1010B
09/17/2021	Water	M3706-12	A Flash Point	Cool 4 deg C	PSEG03	H41	41-43	09/10/2021	1010B

Date/Time 09/13/21 09:55
 Raw Sample Received by: KS(WC)
 Raw Sample Relinquished by: JW(C)

Date/Time 09/13/21 12:45
 Raw Sample Received by: JW(C)
 Raw Sample Relinquished by: KS(WC)