

UB116273

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : PA Instrument ID : Konelab

9/13/2021 13:22

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	93.546	0.0	0.071	
ICB1	2.484	0.0	0.001	
CCV1	237.048	0.0	0.182	
CCB1	2.050	0.0	0.001	
PB139046BS	95.868	0.0	0.073	
M3705-25	327.443	0.0	0.251	
PB139046BL	1.783	0.0	0.001	
CCV2	244.007	0.0	0.187	
CCB2	2.013	0.0	0.001	

N	9
Mean	111.805
SD	126.6918
CV%	113.32

Aquakem v. 7.2AQ1

Results from time period:

Mon Sep 13 12:56:09 2021

Mon Sep 13 13:21:33 2021

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.9606	µg/l	9/13/2021 10:32:10	
5.0PPBCN	A	Total CN	P	5.6913	µg/l	9/13/2021 10:32:11	
10PPBCN	A	Total CN	P	10.5393	µg/l	9/13/2021 10:32:12	
50PPBCN	A	Total CN	P	52.4711	µg/l	9/13/2021 10:32:13	
100PPBCN	A	Total CN	P	98.6147	µg/l	9/13/2021 10:32:14	
250PPBCN	A	Total CN	P	247.2641	µg/l	9/13/2021 10:32:15	
500PPBCN	A	Total CN	P	501.3802	µg/l	9/13/2021 10:32:16	
ICV1	S	Total CN	P	93.5459	µg/l	9/13/2021 12:56:10	
ICB1	S	Total CN	P	2.4842	µg/l	9/13/2021 12:56:12	
CCV1	S	Total CN	P	237.0483	µg/l	9/13/2021 12:56:14	
CCB1	S	Total CN	P	2.0505	µg/l	9/13/2021 12:56:17	
PB139046BS	S	Total CN	P	95.8678	µg/l	9/13/2021 13:02:20	
M3705-25	S	Total CN	P	327.4427	µg/l	9/13/2021 13:02:23	
PB139046BL	S	Total CN	P	1.783	µg/l	9/13/2021 13:21:28	
CCV2	S	Total CN	P	244.0073	µg/l	9/13/2021 13:21:31	
CCB2	S	Total CN	P	2.0133	µg/l	9/13/2021 13:21:33	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : ph

Instrument ID : Konelab

9/13/2021 10:33


Test Total CN

Accepted 9/13/2021 10:33

Factor 1300
Bias -0.001

Coeff. of det. 0.999907

Errors



	Calibrator	Response	Calc. con.	Conc.	Rel. Errors
1	0.0PPBCN	-0.001	-0.9606	0.0000	
2	5.0PPBCN	0.004	5.6913	5.0000	13.8
3	10PPBCN	0.007	10.5393	10.0000	5.4
4	50PPBCN	0.040	52.4711	50.0000	4.9
5	100PPBCN	0.075	98.6147	100.0000	-1.4
6	250PPBCN	0.190	247.2641	250.0000	-1.1
7	500PPBCN	0.385	501.3802	500.0000	0.3

AL
9/13/21