


LB 116274

7471B

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 1.8360e-004

C= 1.1989e-002

Rho= 0.9999346

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.022	0.022	55	0.000	55					1
0.2	0.200	0.200	0.000	1026	0.0 %	1026					0
2.5	2.500	2.531	0.031	13721	0.0 %	13721					1
5.0	5.000	4.911	-0.089	26684	0.0 %	26684					2
7.5	7.500	7.498	-0.002	40775	0.0 %	40775					0
10.0	10.000	10.037	0.037	54600	0.0 %	54600					0

LB116274 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 13 Sep 2021 15:27:50					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	55	-	0.00007471B	PPB		13 Sep 2021 15:28:45	S	Std
0.2 - 1	50.2	1026	-	0.20007471B	PPB		13 Sep 2021 15:31:02	S	Std
2.5 - 1	52.5	13721	-	2.50007471B	PPB		13 Sep 2021 15:33:19	S	Std
5.0 - 1	55	26684	-	5.00007471B	PPB		13 Sep 2021 15:35:36	S	Std
7.5 - 1	57.5	40775	-	7.50007471B	PPB		13 Sep 2021 15:37:52	S	Std
10.0 - 1	510	54600	-	10.00007471B	PPB		13 Sep 2021 15:43:04	S	Std
ICV89 - 1	ICV89	20833	3.8370	-7471B	PPB		13 Sep 2021 15:45:42	U	SMPL
ICB89 - 1	ICB89	-13	0.0096	-7471B	PPB		13 Sep 2021 15:47:57	U	SMPL
CCV27 - 1	CCV27	26120	4.8077	-7471B	PPB		13 Sep 2021 15:50:12	U	SMPL
CCB27 - 1	CCB27	-42	0.0043	-7471B	PPB		13 Sep 2021 15:52:27	U	SMPL
CRA - 1	CRA	1008	0.1971	-7471B	PPB		13 Sep 2021 15:54:42	U	SMPL
HighStd - 1	HighStd	54873	10.0868	-7471B	PPB		13 Sep 2021 15:56:57	U	SMPL
ChkStd - 1	ChkStd	38892	7.1527	-7471B	PPB		13 Sep 2021 15:59:12	U	SMPL
PB139052BL - 1	PBS	-57	0.0015	-7471B	PPB		13 Sep 2021 16:01:30	U	SMPL
PB139052BS - 1	LCSS	20963	3.8609	-7471B	PPB		13 Sep 2021 16:03:46	U	SMPL
M3684-01 - 1	A018015	7351	1.3617	-7471B	PPB		13 Sep 2021 16:06:03	U	SMPL
M3684-02 - 1	A018145	838	0.1658	-7471B	PPB		13 Sep 2021 16:08:19	U	SMPL
M3684-03 - 1	DUP-210908A	3161	0.5924	-7471B	PPB		13 Sep 2021 16:10:36	U	SMPL
M3684-07 - 1	A019015	4444	0.8279	-7471B	PPB		13 Sep 2021 16:12:53	U	SMPL
M3684-08 - 1	A019145	1338	0.2576	-7471B	PPB		13 Sep 2021 16:15:10	U	SMPL
CCV28 - 1	CCV28	26013	4.7880	-7471B	PPB		13 Sep 2021 16:17:24	U	SMPL
CCB28 - 1	CCB28	-40	0.0046	-7471B	PPB		13 Sep 2021 16:19:39	U	SMPL
M3684-09 - 1	F002050	648	0.1310	-7471B	PPB		13 Sep 2021 16:21:55	U	SMPL
M3684-10 - 1	A017015	3595	0.6720	-7471B	PPB		13 Sep 2021 16:24:11	U	SMPL
M3684-11 - 1	A017145	2847	0.5347	-7471B	PPB		13 Sep 2021 16:26:25	U	SMPL
M3684-14 - 1	A008015	3320	0.6215	-7471B	PPB		13 Sep 2021 16:28:41	U	SMPL
M3684-15 - 1	A008145	1854	0.3524	-7471B	PPB		13 Sep 2021 16:30:57	U	SMPL
M3684-16 - 1	A007015	7679	1.4219	-7471B	PPB		13 Sep 2021 16:33:13	U	SMPL
M3684-17 - 1	A007145	349	0.0761	-7471B	PPB		13 Sep 2021 16:35:29	U	SMPL
M3684-18 - 1	A001015	3074	0.5764	-7471B	PPB		13 Sep 2021 16:37:46	U	SMPL
M3684-19 - 1	A001145	405	0.0863	-7471B	PPB		13 Sep 2021 16:40:03	U	SMPL
M3684-19DUP - 1	A001145DUP	383	0.0823	-7471B	PPB		13 Sep 2021 16:42:19	U	SMPL
CCV29 - 1	CCV29	26390	4.8573	-7471B	PPB		13 Sep 2021 16:44:34	U	SMPL
CCB29 - 1	CCB29	-22	0.0079	-7471B	PPB		13 Sep 2021 16:46:50	U	SMPL
M3684-19MS - 1	A001145MS	22371	4.1194	-7471B	PPB		13 Sep 2021 16:49:05	U	SMPL
M3684-19MSD - 1	A001145MSD	21650	3.9870	-7471B	PPB		13 Sep 2021 16:53:59	U	SMPL
M3684-20 - 1	A006015	5580	1.0365	-7471B	PPB		13 Sep 2021 16:56:14	U	SMPL
M3684-21 - 1	A006070	2837	0.5329	-7471B	PPB		13 Sep 2021 16:58:29	U	SMPL
PB139053BL - 1	PBS	49	0.0210	-7471B	PPB		13 Sep 2021 17:00:45	U	SMPL
PB139053BS - 1	LCSS	22736	4.1864	-7471B	PPB		13 Sep 2021 17:03:01	U	SMPL
M3719-03 - 1	SU-04-091021	2412	0.4548	-7471B	PPB		13 Sep 2021 17:05:17	U	SMPL
M3719-03DUP - 1	SU-04-091021DUP	2439	0.4598	-7471B	PPB		13 Sep 2021 17:07:34	U	SMPL
M3719-03MS - 1	SU-04-091021MS	22923	4.2207	-7471B	PPB		13 Sep 2021 17:09:51	U	SMPL
M3719-03MSD - 1	SU-04-091021MSD	22850	4.2073	-7471B	PPB		13 Sep 2021 17:12:07	U	SMPL
CCV30 - 1	CCV30	27039	4.9764	-7471B	PPB		13 Sep 2021 17:14:22	U	SMPL
CCB30 - 1	CCB30	-57	0.0015	-7471B	PPB		13 Sep 2021 17:16:38	U	SMPL
M3685-01 - 1	327	25902	4.7677	-7471B	PPB		13 Sep 2021 17:18:53	U	SMPL
M3691-01 - 1	OK-03-090921	5667	1.0525	-7471B	PPB		13 Sep 2021 17:21:08	U	SMPL
M3692-01 - 1	WM-1	2251	0.4253	-7471B	PPB		13 Sep 2021 17:23:24	U	SMPL
M3693-01 - 1	WFM-1	1324	0.2551	-7471B	PPB		13 Sep 2021 17:25:40	U	SMPL
M3707-01 - 1	PL-02-091021	1429	0.2744	-7471B	PPB		13 Sep 2021 17:27:56	U	SMPL
M3711-01 - 1	TR-04-091021	1725	0.3287	-7471B	PPB		13 Sep 2021 17:30:12	U	SMPL
M3715-01 - 1	CL-2-091021	2011	0.3812	-7471B	PPB		13 Sep 2021 17:32:28	U	SMPL
M3718-01 - 1	OR-03-091021	4461	0.8310	-7471B	PPB		13 Sep 2021 17:34:45	U	SMPL
M3719-01 - 1	SU-01-091021	2660	0.5004	-7471B	PPB		13 Sep 2021 17:37:02	U	SMPL
CCV31 - 1	CCV31	26354	4.8507	-7471B	PPB		13 Sep 2021 17:39:18	U	SMPL
CCB31 - 1	CCB31	-38	0.0050	-7471B	PPB		13 Sep 2021 17:41:34	U	SMPL
M3684-19LX5 - 1		62	0.0234	-7471B	PPB		13 Sep 2021 17:43:49	U	SMPL
M3684-19A - 1		22130	4.0751	-7471B	PPB		13 Sep 2021 17:46:05	U	SMPL
M3719-03LX5 - 1		388	0.0832	-7471B	PPB		13 Sep 2021 17:48:20	U	SMPL
M3719-03A - 1		23254	4.2815	-7471B	PPB		13 Sep 2021 17:50:35	U	SMPL
CCV32 - 1	CCV32	26718	4.9175	-7471B	PPB		13 Sep 2021 17:52:51	U	SMPL
CCB32 - 1	CCB32	-47	0.0034	-7471B	PPB		13 Sep 2021 17:55:06	U	SMPL