



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116276

Slope : 98.7

BalanceID: Wc Sc-04

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	09/13/2021	17:26
2	CAL2	1	Water	NA	NA	20.6	7.00	09/13/2021	17:27
3	CAL3	1	Water	NA	NA	20.6	10.03	09/13/2021	17:30
4	ICV	1	Water	NA	NA	20.4	7.00	09/13/2021	17:32
5	CCV1	1	Water	NA	NA	20.6	2.01	09/13/2021	17:34
6	M3722-02	1	Solid	20.03	20	22.3	10.10	09/13/2021	17:36
7	M3722-02DUP	1	Solid	20.04	20	22.4	10.11	09/13/2021	17:37
8	M3722-05	1	Solid	20.02	20	21.9	9.97	09/13/2021	17:40
9	M3724-02	1	Solid	20.03	20	23.0	6.82	09/13/2021	17:42
10	M3724-04	1	Solid	20.04	20	23.4	7.18	09/13/2021	17:45
11	CCV2	1	Water	NA	NA	20.6	12.02	09/13/2021	17:46

WORKLIST(Hardcopy Internal Chain)

LB 116276

WorkList Name : CORROSIVITY M3722

WorkList ID : 152200

Department : Wet-Chemistry

Date : 09-13-2021 16:17:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/13/2021	Solid	M3722-02	Corrosivity	Cool 4 deg C	PSEG03	H41	WALLS	09/13/2021	9045D
09/13/2021	Solid	M3722-05	Corrosivity	Cool 4 deg C	PSEG03	H41	FLOOR	09/13/2021	9045D
09/13/2021	Solid	M3724-02	Corrosivity	Cool 4 deg C	PSEG03	H41	31190	09/13/2021	9045D
09/13/2021	Solid	M3724-04	Corrosivity	Cool 4 deg C	PSEG03	H41	31191	09/13/2021	9045D

Date/Time 09-13-21 16:20
 Raw Sample Received by: JP (WLC)
 Raw Sample Relinquished by: JTC (sm)

Date/Time 09-13-21 18:00
 Raw Sample Received by: JTC (sm)
 Raw Sample Relinquished by: JP (WLC)