



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116282

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	09/14/2021	09:35
2	M3724-05		1	10	100.00	>212.0	>212.0	09/14/2021	09:58
3	M3724-05DUP		1	10	100.00	>212.0	>212.0	09/14/2021	10:20
4	M3724-06		1	12	100.00	>212.0	>212.0	09/14/2021	10:45
5	M3724-07		1	10	100.00	>212.0	>212.0	09/14/2021	11:10
6	M3724-08		1	11	100.00	>212.0	>212.0	09/14/2021	11:35
7	M3724-09		1	12	100.00	>212.0	>212.0	09/14/2021	12:05
8	M3724-21		1	10	100.00	>212.0	>212.0	09/14/2021	12:40
9	M3724-22		1	12	100.00	>212.0	>212.0	09/14/2021	13:05
10	M3724-23		1	11	100.00	>212.0	>212.0	09/14/2021	13:32

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB116282

WorkList Name : fp91421 WorkList ID : 15221 Department : Wet-Chemistry Date : 09-14-2021 09:00:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method	
09/16/2021	Water	M3724-05	C	Flash Point	Cool 4 deg C	PSEG03	H41	31192	09/13/2021	1010B
09/16/2021	Water	M3724-06	C	Flash Point	Cool 4 deg C	PSEG03	H41	31193	09/13/2021	1010B
09/16/2021	Water	M3724-07	C	Flash Point	Cool 4 deg C	PSEG03	H41	31194	09/13/2021	1010B
09/16/2021	Water	M3724-08	C	Flash Point	Cool 4 deg C	PSEG03	H41	31195	09/13/2021	1010B
09/16/2021	Water	M3724-09	C	Flash Point	Cool 4 deg C	PSEG03	H41	31196	09/13/2021	1010B
09/16/2021	Water	M3724-21	A	Flash Point	Cool 4 deg C	PSEG03	H41	27-28-29	09/13/2021	1010B
09/16/2021	Water	M3724-22	A	Flash Point	Cool 4 deg C	PSEG03	H41	30-31-32	09/13/2021	1010B
09/16/2021	Water	M3724-23	A	Flash Point	Cool 4 deg C	PSEG03	H41	33-34-35-36	09/13/2021	1010B

Date/Time 09/14/21 09:05
 Raw Sample Received by: KS(WC)
 Raw Sample Relinquished by: MB (DY)

Date/Time 09/14/21 13:35
 Raw Sample Received by: MB (WC)
 Raw Sample Relinquished by: KS(WC)