



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: ASTM Oil and Grease
Run Number: LB116285
Analysis Date: 09/14/2021
BalanceID: Wc Sc-06
OvenID: Ext Oven-03

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 09/14/2021
Extraction IN Time: 11:30
Extraction OUT Time: 12:15
Thermometer ID: Ext Oven#03

Dish #	Lab ID	Client ID	Matrix	pH	Sample Weight (g)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB116285BL	LB116285BL	WATER	1.3	1000	100	3.0147	3.0147	0	3.0148	3.0148	0.0001	0.1
2	LB116285BS	LB116285BS	WATER	1.3	1000	100	3.0579	3.0579	0	3.0741	3.0741	0.0162	16.2
3	M3722-03	WALLS	WATER	1.6	1000	100	2.9951	2.9951	0	2.9957	2.9957	0.0006	0.6
4	M3722-03DUP	WALLSDUP	WATER	1.6	1000	100	3.1563	3.1563	0	3.1568	3.1568	0.0005	0.5
5	M3722-03MS	WALLS	WATER	1.6	1000	100	3.0654	3.0654	0	3.0859	3.0859	0.0205	20.5
6	M3722-03MSD	WALLS	WATER	1.6	1000	100	2.9841	2.9841	0	3.0037	3.0037	0.0196	19.6
7	M3722-06	FLOOR	WATER	1.6	1000	100	3.0494	3.0494	0	3.0501	3.0501	0.0007	0.7



QC Batch# LB116285 Test: ASTM Oil and Grease Analysis Date: 09/14/2021

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2870
pH Paper 0-14	M4909
Sodium Sulfate	EP2179
1:1 HCL	WP92935
Silica Gel	NA
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	2.5 ML	WP92936
LCSWD	NA	NA
MS/MSD	2.5 ML	WP92937

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance:	0.0018	(0.0018-0.0022)	In OVEN TEMP1 :	70 °C	Dessicator Time In1	13:26
1.0000 gram Balance:	1.0004	(0.9950-1.0050)	In Time1:	12:40		
Bal Check Time:	11:40		Out OVEN TEMP1:	70 °C	Dessicator Time Out1	14:00
			Out Time1	13:25		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	(0.0018-0.0022)	In OVEN TEMP2 :	<u>71 °C</u>	Dessicator Time In2	<u>15:01</u>
1.0000 gram Balance:	<u>1.0005</u>	(0.9950-1.0050)	In Time2:	<u>14:30</u>		
Bal Check Time:	<u>16:00</u>		Out OVEN TEMP2:	<u>71 °C</u>	Dessicator Time Out2	<u>15:30</u>
			Out Time2:	<u>15:00</u>		

WORKLIST(Hardcopy Internal Chain)

18116285

WorkList Name : astm oil & grease m3722 WorkList ID : 152233 Department : Wet-Chemistry Date : 09-14-2021 11:15:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/17/2021	Solid	M3722-03	ASTM Oil and Grease	Cool 4 deg C	PSEG03	H41	WALLS	09/13/2021	1664A
09/17/2021	Solid	M3722-06	ASTM Oil and Grease	Cool 4 deg C	PSEG03	H41	FLOOR	09/13/2021	1664A

Date/Time 09-14-21 11:20
 Raw Sample Received by: JP(WC)
 Raw Sample Relinquished by: JP(WC)

Date/Time 09-14-21 13:00
 Raw Sample Received by: JP(WC)
 Raw Sample Relinquished by: JP(WC)