



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB116293
Analysis Date: 09/14/2021
BalanceID: Wc Sc-06
OvenID: Ext Oven03

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 09/14/2021
Extraction IN Time: 13:30
Extraction OUT Time: 14:15
Thermometer ID: Ext Oven#03

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB116293BL	LB116293BL	WATER	1.3	1000	100	3.0534	3.0534	0	3.0535	3.0535	0.0001	0.1
2	LB116293BS	LB116293BS	WATER	1.3	1000	100	3.1671	3.1671	0	3.1834	3.1834	0.0163	16.3
3	M3732-02	DRAIN-WATER-TANK-1	WATER	1.6	1000	100	3.0523	3.0523	0	3.0642	3.0642	0.0119	11.9
4	M3732-03	M3732-02MS	WATER	1.6	1000	100	3.0143	3.0143	0	3.0455	3.0455	0.0312	31.2
5	M3732-04	M3732-02MSD	WATER	1.6	1000	100	3.1152	3.1152	0	3.1463	3.1463	0.0311	31.1



Chemical Name	Chemical Lot #
HEXANE	W2870
pH Paper 0-14	M4909
Sodium Sulfate	EP2179
1:1 HCL	WP92935
Silica Gel	W2839
Sand	NA

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP92936
LCSWD	NA	NA
MS/MSD	5.00 ML	WP92937

Analytical Balance ID # : WC SC-6

0.0020 gram Balance:	<u>0.0018</u>	(0.0018-0.0022)	In OVEN TEMP1 :	<u>70 °C</u>	Dessicator Time In1	<u>15:21</u>
1.0000 gram Balance:	<u>1.0004</u>	(0.9950-1.0050)	In Time1:	<u>14:40</u>		
Bal Check Time:	<u>13:40</u>		Out OVEN TEMP1:	<u>70 °C</u>	Dessicator Time Out1	<u>16:00</u>
			Out Time1	<u>15:20</u>		

0.0020 gram Balance:	<u>0.0021</u>	(0.0018-0.0022)	In OVEN TEMP2 :	<u>71 °C</u>	Dessicator Time In2	<u>17:01</u>
1.0000 gram Balance:	<u>1.0005</u>	(0.9950-1.0050)	In Time2:	<u>16:30</u>		
Bal Check Time:	<u>18:00</u>		Out OVEN TEMP2:	<u>71 °C</u>	Dessicator Time Out2	<u>17:35</u>
			Out Time2:	<u>17:00</u>		

WORKLIST(Hardcopy Internal Chain)

116293

WorkList Name : TPH M3717

Department : Wet-Chemistry

WorkList ID : 152235

Date : 09-14-2021 13:18:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/21/2021	Water	M3732-02	TPH	Conc H2SO4 tc	MAJO01	D11	DRAIN-WATER-TANK-1	09/14/2021	1664A
09/21/2021	Water	M3732-03	TPH	Conc H2SO4 tc	MAJO01	D11	M3732-02MS	09/14/2021	1664A
09/21/2021	Water	M3732-04	TPH	Conc H2SO4 tc	MAJO01	D11	M3732-02MSD	09/14/2021	1664A

13:20

Date/Time 09-14-21 13:20
Raw Sample Received by: JP (WLC)
Raw Sample Relinquished by: CP SM

14:00

Date/Time 09-14-21 14:00
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JP (WLC)