



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: jignesh

Date: 09/15/2021

OVEN IN TEMP1: 103 °C

OVEN OUT TEMP1: 104 °C

DATE & TIME IN1: 09/14/2021 16:00 DATE & TIME OUT1: 09/15/2021 08:00

OVEN IN TEMP2: 104 °C

OVEN OUT TEMP2: 104 °C

DATE & TIME IN2: 09/15/2021 08:30 DATE & TIME OUT2: 09/15/2021 10:00

Run Number: LB116294

BalanceID: Wc SC-06

OvenID: Wc Oven-01

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Dish + Sample Weight (g)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result %
1	LB116294BL	LB116294BL	68.7413	68.7413	68.7413	68.7413	68.7413	68.7413	0.0000	0
2	M3730-04	WALL-BRICK-FLOOR	55.8888	55.8888	75.0429	74.5818	74.5818	74.5818	18.6930	97.6
3	M3730-04DUP	WALL-BRICK-FLOORDUP	59.9825	59.9825	81.9952	81.6022	81.6022	81.6022	21.6197	98.2

A = Final Empty Dish Weight (g)

B = Dish + Sample Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result \%} = (C - A) * 100 / (B - A)$$

2B 116294

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ts m3730 s WorkList ID : 152239 Department : Wet-Chemistry Date : 09-14-2021 14:12:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/27/2021	Solid	M3730-04	TS	Cool 4 deg C	PSEG03	H41	WALL-BRICK-FLOOR	09/14/2021	SM2540 B

Date/Time 09-14-21 14:20
Raw Sample Received by: JP LWC
Raw Sample Relinquished by: CP SM

Date/Time 09-14-21 17:00
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JP LWC