



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116296

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/14/2021	14:30
2	CAL2	1	Water	NA	NA	20.3	7.00	09/14/2021	14:31
3	CAL3	1	Water	NA	NA	20.3	10.03	09/14/2021	14:32
4	ICV	1	Water	NA	NA	20.4	6.99	09/14/2021	14:33
5	CCV1	1	Water	NA	NA	20.3	2.01	09/14/2021	14:35
6	M3728-01	1	Water	NA	NA	20.6	8.90	09/14/2021	14:50
7	M3732-05	1	Water	NA	NA	20.4	6.16	09/14/2021	14:52
8	M3732-05DUP	1	Water	NA	NA	20.5	6.17	09/14/2021	14:53
9	CCV2	1	Water	NA	NA	20.4	12.02	09/14/2021	15:00

WORKLIST(Hardcopy Internal Chain)

YB 116296

WorkList Name : ph m3728 w WorkList ID : 152244 Department : Wet-Chemistry Date : 09-14-2021 14:43:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/14/2021	Water	M3728-01	L pH	Cool 4 deg C	PSEG03	H51	N42049	09/14/2021	9040C
09/14/2021	Water	M3732-05	✓ pH	Cool 4 deg C	MAJO01	D11	DRAIN-WATER-TANK-1	09/14/2021	9040C

Date/Time 09-14-21 14:45
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: OP SM

Date/Time 09-14-21 14:10
 Raw Sample Received by: OP SM
 Raw Sample Relinquished by: JP (WC)