



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116297

Slope : 98.7

BalanceID: Wc Sc-06

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/14/2021	14:30
2	CAL2	1	Water	NA	NA	20.3	7.00	09/14/2021	14:31
3	CAL3	1	Water	NA	NA	20.3	10.03	09/14/2021	14:32
4	ICV	1	Water	NA	NA	20.4	6.99	09/14/2021	14:33
5	CCV1	1	Water	NA	NA	20.3	2.01	09/14/2021	14:35
6	M3730-05	1	Solid	20.03	20	22.7	10.04	09/14/2021	14:40
7	M3730-05DUP	1	Solid	20.04	20	22.8	10.03	09/14/2021	14:41
8	CCV2	1	Water	NA	NA	20.4	12.03	09/14/2021	14:45

WORKLIST(Hardcopy Internal Chain)

116297

WorkList Name: CORROSIONITY M3730 WorkList ID: 152238 Department: Wet-Chemistry Date: 09-14-2021 14:12:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/14/2021	Solid	M3730-05	Corrosivity	Cool 4 deg C	PSEG03	H41	WALL-BRICK-FLOOR	09/14/2021	9045D

Date/Time 09-14-21 14:20
 Raw Sample Received by: JP LWC
 Raw Sample Relinquished by: JP LWC

Date/Time 09-14-21 14:00
 Raw Sample Received by: JP LWC
 Raw Sample Relinquished by: JP LWC