


LB 116301

SFAM01.1

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 2.2990e-004 slope

C= -3.5487e-003 intercept

Rho= 0.9999041

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.010	0.010	57	0.000	57					-
0.05	0.050	0.044	-0.006	205	0.0 %	205					-12
0.20	0.200	0.187	-0.013	827	0.0 %	827					-7
2.50	2.500	2.469	-0.031	10755	0.0 %	10755					-1
5.00	5.000	5.019	0.019	21846	0.0 %	21846					0
7.50	7.500	7.602	0.102	33083	0.0 %	33083					1
10.0	10.000	9.920	-0.080	43164	0.0 %	43164					-1

LB116301

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	57	-		0 SFAM01.1	PPB	9/14/2021 16:02	Std
	0.05 S01	205	-	0.05	SFAM01.1	PPB	9/14/2021 16:05	Std
	0.2 S02	827	-	0.2	SFAM01.1	PPB	9/14/2021 16:07	Std
	2.5 S03	10755	-	2.5	SFAM01.1	PPB	9/14/2021 16:09	Std
	5 S04	21846	-	5	SFAM01.1	PPB	9/14/2021 16:12	Std
	7.5 S05	33083	-	7.5	SFAM01.1	PPB	9/14/2021 16:14	Std
	10 S06	43164	-	10	SFAM01.1	PPB	9/14/2021 16:16	Std
ICV070	ICV070	17197	3.9501 -		SFAM01.1	PPB	9/14/2021 16:19	SMPL
ICB070	ICB070	-29	-0.0102 -		SFAM01.1	PPB	9/14/2021 16:21	SMPL
CCV036	CCV036	21568	4.955 -		SFAM01.1	PPB	9/14/2021 16:23	SMPL
CCB036	CCB036	-42	-0.0132 -		SFAM01.1	PPB	9/14/2021 16:26	SMPL
PB139094BL	PBW094	46	0.007 -		SFAM01.1	PPB	9/14/2021 16:28	SMPL
M3710-01	EW130	60	0.0102 -		SFAM01.1	PPB	9/14/2021 16:30	SMPL
M3710-02	EW131	36	0.0047 -		SFAM01.1	PPB	9/14/2021 16:32	SMPL
M3710-03	EW132	41	0.0059 -		SFAM01.1	PPB	9/14/2021 16:35	SMPL
M3710-04	EW132D	51	0.0082 -		SFAM01.1	PPB	9/14/2021 16:37	SMPL
M3710-05	EW132S	4894	1.1216 -		SFAM01.1	PPB	9/14/2021 16:39	SMPL
CCV037	CCV037	21361	4.9074 -		SFAM01.1	PPB	9/14/2021 16:41	SMPL
CCB037	CCB037	-22	-0.0086 -		SFAM01.1	PPB	9/14/2021 16:44	SMPL