



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 9/15/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:30
In Date: 09/14/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 09/15/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB116304

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M3730-01	FLOOR	1	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3730-02	WALLS	2	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3730-03	BRICKS	3	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3730-04	WALL-BRICK-FLOOR	4	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3731-01	HR-02-091421	5	1.16	9.86	9.13	91.6	
M3731-02	HR-02-091421-EPH-2	6	1.16	9.57	8.97	92.9	
M3731-02D	HR-02-091421-EPH-2DUP	7	1.15	9.59	8.99	92.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-091421

WorkList ID : 152211

Department : Wet-Chemistry

Date : 09-14-2021 08:42:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/24/2021	Solid	M3730-01	Percent Solids	Cool 4 deg C	PSEG03	H41	FLOOR	09/14/2021	Chemtech -SO
09/24/2021	Solid	M3730-02	Percent Solids	Cool 4 deg C	PSEG03	H41	WALLS	09/14/2021	Chemtech -SO
09/24/2021	Solid	M3730-03	Percent Solids	Cool 4 deg C	PSEG03	H41	BRICKS	09/14/2021	Chemtech -SO
09/24/2021	Solid	M3730-04	Percent Solids	Cool 4 deg C	PSEG03	H41	WALL-BRICK-FLOOR	09/14/2021	Chemtech -SO
09/21/2021	Solid	M3731-01	Percent Solids	Cool 4 deg C	PSEG05	D51	HR-02-091421	09/14/2021	Chemtech -SO
09/21/2021	Solid	M3731-02	Percent Solids	Cool 4 deg C	PSEG05	D51	HR-02-091421-EPH-2	09/14/2021	Chemtech -SO

Date/Time 09-14-21 16:00

Raw Sample Received by: JRC

Raw Sample Relinquished by: KS (WC)

Date/Time 09-14-21 16:45

Raw Sample Received by: KS (WC)

Raw Sample Relinquished by: JRC