


LB116310

7471B

INSTRUMENT ID: CV4



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/oid
0.0	0.000	-0.037	-0.037	85	0.000	85					-13
0.2	0.200	0.175	-0.025	1177	0.0 %	1177					3
2.5	2.500	2.582	0.082	13576	0.0 %	13576					-1
5.0	5.000	4.948	-0.052	25763	0.0 %	25763					2
7.5	7.500	7.616	0.116	39506	0.0 %	39506					-1
10.0	10.000	9.915	-0.085	51346	0.0 %	51346					

LB116310 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Sep 2021 09:20:49

Sample ID	Extended ID	µ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	85	-	0.00007471B	PPB	15 Sep 2021 09:23:53	S	Std
0.2 - 1	50.2	1177	-	0.20007471B	PPB	15 Sep 2021 09:26:09	S	Std
2.5 - 1	52.5	13576	-	2.50007471B	PPB	15 Sep 2021 09:28:26	S	Std
5.0 - 1	55	25763	-	5.00007471B	PPB	15 Sep 2021 09:30:43	S	Std
7.5 - 1	57.5	39506	-	7.50007471B	PPB	15 Sep 2021 09:32:59	S	Std
10.0 - 1	510	51346	-	10.00007471B	PPB	15 Sep 2021 09:37:51	S	Std
ICV91 - 1	ICV91	21496	4.1198	-7471B	PPB	15 Sep 2021 09:40:32	U	SMPL
ICB91 - 1	ICB91	117	-0.0309	-7471B	PPB	15 Sep 2021 09:42:47	U	SMPL
CCV38 - 1	CCV38	25367	4.8714	-7471B	PPB	15 Sep 2021 09:45:02	U	SMPL
CCB38 - 1	CCB38	111	-0.0320	-7471B	PPB	15 Sep 2021 09:47:17	U	SMPL
CRA - 1	CRA	1151	0.1699	-7471B	PPB	15 Sep 2021 09:49:32	U	SMPL
HighStd - 1	HighStd	52277	10.0960	-7471B	PPB	15 Sep 2021 09:51:47	U	SMPL
ChkStd - 1	ChkStd	37190	7.1668	-7471B	PPB	15 Sep 2021 09:54:02	U	SMPL
PB139109BL - 1	PBS	131	-0.0282	-7471B	PPB	15 Sep 2021 09:58:52	U	SMPL
PB139109BS - 1	LCSS	20915	4.0070	-7471B	PPB	15 Sep 2021 10:01:08	U	SMPL
M3687-17 - 1	A009015	1901	0.3155	-7471B	PPB	15 Sep 2021 10:03:24	U	SMPL
M3687-18 - 1	A009095	652	0.0730	-7471B	PPB	15 Sep 2021 10:05:41	U	SMPL
M3687-19 - 1	A016015	2788	0.4877	-7471B	PPB	15 Sep 2021 10:07:58	U	SMPL
M3687-20 - 1	A016095	1553	0.2479	-7471B	PPB	15 Sep 2021 10:10:15	U	SMPL
M3687-21 - 1	A005015	43889	8.4674	-7471B	PPB	15 Sep 2021 10:12:30	U	SMPL
CCV39 - 1	CCV39	25453	4.8881	-7471B	PPB	15 Sep 2021 10:14:45	U	SMPL
CCB39 - 1	CCB39	125	-0.0293	-7471B	PPB	15 Sep 2021 10:17:02	U	SMPL
M3687-22 - 1	A005095	2251	0.3834	-7471B	PPB	15 Sep 2021 10:19:17	U	SMPL
M3687-23 - 1	A010015	3456	0.6174	-7471B	PPB	15 Sep 2021 10:21:32	U	SMPL
M3687-24 - 1	A010145	1302	0.1992	-7471B	PPB	15 Sep 2021 10:23:47	U	SMPL
M3687-25 - 1	A015015	3923	0.7081	-7471B	PPB	15 Sep 2021 10:26:03	U	SMPL
M3687-26 - 1	A015095	1405	0.2192	-7471B	PPB	15 Sep 2021 10:28:19	U	SMPL
M3687-27 - 1	A011145	1131	0.1660	-7471B	PPB	15 Sep 2021 10:30:35	U	SMPL
M3687-28 - 1	A004015	3260	0.5793	-7471B	PPB	15 Sep 2021 10:32:51	U	SMPL
M3687-29 - 1	A004145	1292	0.1972	-7471B	PPB	15 Sep 2021 10:35:08	U	SMPL
M3687-30 - 1	A003015	2801	0.4902	-7471B	PPB	15 Sep 2021 10:37:25	U	SMPL
M3687-31 - 1	A003145	329	0.0103	-7471B	PPB	15 Sep 2021 10:39:41	U	SMPL
CCV40 - 1	CCV40	25709	4.9378	-7471B	PPB	15 Sep 2021 10:41:56	U	SMPL
CCB40 - 1	CCB40	143	-0.0258	-7471B	PPB	15 Sep 2021 10:44:11	U	SMPL
M3687-32 - 1	A013015	5042	0.9253	-7471B	PPB	15 Sep 2021 10:46:27	U	SMPL
M3687-33 - 1	A013115	1507	0.2390	-7471B	PPB	15 Sep 2021 10:48:43	U	SMPL
M3687-34 - 1	A012145	394	0.0229	-7471B	PPB	15 Sep 2021 10:50:58	U	SMPL
M3687-35 - 1	A002015	2610	0.4531	-7471B	PPB	15 Sep 2021 10:53:14	U	SMPL
M3687-36 - 1	A002145	494	0.0423	-7471B	PPB	15 Sep 2021 10:55:29	U	SMPL
M3687-36DUP - 1	A002145DUP	389	0.0219	-7471B	PPB	15 Sep 2021 10:57:45	U	SMPL
M3687-36MS - 1	A002145MS	20609	3.9476	-7471B	PPB	15 Sep 2021 11:00:01	U	SMPL
M3687-36MSD - 1	A002145MSD	22786	4.3703	-7471B	PPB	15 Sep 2021 11:02:18	U	SMPL
PB139110BL - 1	PBS	147	-0.0251	-7471B	PPB	15 Sep 2021 11:04:35	U	SMPL
PB139110BS - 1	LCSS	20029	3.8350	-7471B	PPB	15 Sep 2021 11:06:50	U	SMPL
CCV41 - 1	CCV41	25498	4.8968	-7471B	PPB	15 Sep 2021 11:09:06	U	SMPL
CCB41 - 1	CCB41	222	-0.0105	-7471B	PPB	15 Sep 2021 11:11:21	U	SMPL
M3720-02 - 1	WQ-SOILPILE	11650	2.2082	-7471B	PPB	15 Sep 2021 11:13:38	U	SMPL
M3723-01 - 1	BU-03-091321	5385	0.9919	-7471B	PPB	15 Sep 2021 11:15:53	U	SMPL
M3724-01 - 1	31190	16672	3.1833	-7471B	PPB	15 Sep 2021 11:18:08	U	SMPL
M3724-03 - 1	31191	168	-0.0210	-7471B	PPB	15 Sep 2021 11:20:24	U	SMPL
M3725-01 - 1	HR-01-091321	3863	0.6964	-7471B	PPB	15 Sep 2021 11:27:46	U	SMPL
M3725-01DUP - 1	HR-01-091321DUP	4012	0.7253	-7471B	PPB	15 Sep 2021 11:30:02	U	SMPL
M3725-01MS - 1	HR-01-091321MS	27707	5.3257	-7471B	PPB	15 Sep 2021 11:32:18	U	SMPL
M3725-01MSD - 1	HR-01-091321MSD	23777	4.5627	-7471B	PPB	15 Sep 2021 11:34:35	U	SMPL
M3731-01 - 1	HR-02-091421	1719	0.2801	-7471B	PPB	15 Sep 2021 11:36:52	U	SMPL
CCV42 - 1	CCV42	26466	5.0848	-7471B	PPB	15 Sep 2021 11:39:07	U	SMPL
CCB42 - 1	CCB42	235	-0.0080	-7471B	PPB	15 Sep 2021 11:41:23	U	SMPL
M3687-36LX5 - 1		157	-0.0231	-7471B	PPB	15 Sep 2021 11:43:38	U	SMPL
M3687-36A - 1		21813	4.1814	-7471B	PPB	15 Sep 2021 11:45:54	U	SMPL
M3725-01LX5 - 1		807	0.1031	-7471B	PPB	15 Sep 2021 11:48:10	U	SMPL
M3725-01A - 1		26242	5.0413	-7471B	PPB	15 Sep 2021 11:50:25	U	SMPL
CCV43 - 1	CCV43	25532	4.9034	-7471B	PPB	15 Sep 2021 11:52:41	U	SMPL
CCB43 - 1	CCB43	160	-0.0225	-7471B	PPB	15 Sep 2021 11:54:56	U	SMPL