


LB 116321

SFAM01.1

INSTRUMENT ID: CV1



A= 0.0000e+000

B= 2.4199e-004 slope

C= 7.5883e-003 y-intercept

Rho= 0.9999736

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.00	0.000	0.016	0.016	34	0.000	34					
0.05	0.050					205					
0.20	0.200	0.195	-0.005	774	0.0 %	774					
2.50	2.500	2.476	-0.024	10200	0.0 %	10200					
5.00	5.000	4.990	-0.010	20589	0.0 %	20589					
7.50	7.500	7.551	0.051	31172	0.0 %	31172					
10.0	10.000	9.973	-0.027	41179	0.0 %	41179					

0/0

11

-3

-1

0

1

0

LB116321
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	34	-		0 SFAM01.1	PPB	9/15/2021 13:38	Std
	0.2 S01	774	-		0.2 SFAM01.1	PPB	9/15/2021 13:41	Std
	2.5 S02	10200	-		2.5 SFAM01.1	PPB	9/15/2021 13:43	Std
	5 S03	20589	-		5 SFAM01.1	PPB	9/15/2021 13:45	Std
	7.5 S04	31172	-		7.5 SFAM01.1	PPB	9/15/2021 13:48	Std
	10 S05	41179	-		10 SFAM01.1	PPB	9/15/2021 13:52	Std
ICV071	ICV071	16333	3.96 -		SFAM01.1	PPB	9/15/2021 13:55	SMPL
ICB071	ICB071	-29	0.0006 -		SFAM01.1	PPB	9/15/2021 13:57	SMPL
CCV044	CCV044	20645	5.0035 -		SFAM01.1	PPB	9/15/2021 14:00	SMPL
CCB044	CCB044	-39	-0.0018 -		SFAM01.1	PPB	9/15/2021 14:02	SMPL
PB139095BL	PBW095	18	0.0119 -		SFAM01.1	PPB	9/15/2021 14:04	SMPL
M3713-13	JCF16	25	0.0136 -		SFAM01.1	PPB	9/15/2021 14:06	SMPL
CCV045	CCV045	20753	5.0296 -		SFAM01.1	PPB	9/15/2021 14:09	SMPL
CCB045	CCB045	-39	-0.0018 -		SFAM01.1	PPB	9/15/2021 14:11	SMPL