



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116332

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	09/15/2021	16:25
2	CAL2	1	Water	NA	NA	21.1	7.01	09/15/2021	16:26
3	CAL3	1	Water	NA	NA	21.1	10.04	09/15/2021	16:27
4	ICV	1	Water	NA	NA	21.2	7.00	09/15/2021	16:30
5	CCV1	1	Water	NA	NA	21.2	2.01	09/15/2021	16:32
6	M3705-20	1	Water	NA	NA	23.0	9.24	09/15/2021	16:35
7	CCV2	1	Water	NA	NA	21.4	12.02	09/15/2021	17:01