



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116335

Slope : 99.3

BalanceID: Wc Sc-04

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	09/15/2021	16:25
2	CAL2	1	Water	NA	NA	21.1	7.01	09/15/2021	16:26
3	CAL3	1	Water	NA	NA	21.1	10.04	09/15/2021	16:27
4	ICV	1	Water	NA	NA	21.2	7.00	09/15/2021	16:30
5	CCV1	1	Water	NA	NA	21.2	2.01	09/15/2021	16:32
6	M3757-01	1	Solid	20.03	20	24.3	8.32	09/15/2021	17:25
7	M3757-01DUP	1	Solid	20.04	20	24.4	8.33	09/15/2021	17:26
8	CCV2	1	Water	NA	NA	21.3	12.02	09/15/2021	17:28

116335

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S M3757 WorkList ID : 152278 Department : Wet-Chemistry Date : 09-15-2021 16:42:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/15/2021	Solid	M3757-01	pH	Cool 4 deg C	PSEG05	H52	AU-06-091521	09/15/2021	9045D

Date/Time 09-15-21 16:50
Raw Sample Received by: JR (WC)
Raw Sample Relinquished by: _____

Date/Time 09-15-21 17:50
Raw Sample Received by: _____
Raw Sample Relinquished by: JR (WC)