


LB 116341

SFAM01.1

INSTRUMENT ID: CVI

 μ Abs.:

49331



Linear

A= 0.0000e+000

B= 2.0355e-004 *slope*C= -4.6987e-002 *y-intercept*

Rho= 0.9999830

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.008	0.008	269	0.000	269					=
0.05	0.050					205					6
0.20	0.200	0.211	0.011	1266	0.0 %	1266					0
2.50	2.500	2.488	-0.012	12455	0.0 %	12455					-1
5.00	5.000	4.964	-0.036	24620	0.0 %	24620					0
7.50	7.500	7.534	0.034	37245	0.0 %	37245					0
10.0	10.000	9.994	-0.006	49331	0.0 %	49331					0

LB116341

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	269	-		0 SFAM01.1	PPB	9/16/2021 12:27	Std
	0.2 S01	1266	-	0.2	SFAM01.1	PPB	9/16/2021 12:30	Std
	2.5 S02	12455	-	2.5	SFAM01.1	PPB	9/16/2021 12:32	Std
	5 S03	24620	-	5	SFAM01.1	PPB	9/16/2021 12:34	Std
	7.5 S04	37245	-	7.5	SFAM01.1	PPB	9/16/2021 12:36	Std
	10 S05	49331	-	10	SFAM01.1	PPB	9/16/2021 12:39	Std
ICV072	ICV072	19352	3.8922 -		SFAM01.1	PPB	9/16/2021 12:41	SMPL
ICB072	ICB072	181	-0.0101 -		SFAM01.1	PPB	9/16/2021 12:44	SMPL
CCV051	CCV051	24522	4.9445 -		SFAM01.1	PPB	9/16/2021 12:46	SMPL
CCB051	CCB051	170	-0.0124 -		SFAM01.1	PPB	9/16/2021 12:48	SMPL
PB139147BL	PBS147	255	0.0049 -		SFAM01.1	PPB	9/16/2021 12:53	SMPL
M3713-01	JCF03	2979	0.5594 -		SFAM01.1	PPB	9/16/2021 13:00	SMPL
M3713-02	JCF03D	3112	0.5865 -		SFAM01.1	PPB	9/16/2021 13:02	SMPL
M3713-03	JCF03S	13609	2.7232 -		SFAM01.1	PPB	9/16/2021 13:04	SMPL
M3713-04	JCF04	8828	1.75 -		SFAM01.1	PPB	9/16/2021 13:07	SMPL
M3713-05	JCF05	6115	1.1977 -		SFAM01.1	PPB	9/16/2021 13:09	SMPL
M3713-06	JCF06	1262	0.2099 -		SFAM01.1	PPB	9/16/2021 13:11	SMPL
M3713-07	JCF08	5537	1.0801 -		SFAM01.1	PPB	9/16/2021 13:14	SMPL
M3713-08	JCF11	7295	1.4379 -		SFAM01.1	PPB	9/16/2021 13:16	SMPL
M3713-09	JCF12	7423	1.464 -		SFAM01.1	PPB	9/16/2021 13:18	SMPL
M3713-10	JCF13	11073	2.2069 -		SFAM01.1	PPB	9/16/2021 13:20	SMPL
M3713-11	JCF14	5258	1.0233 -		SFAM01.1	PPB	9/16/2021 13:23	SMPL
M3713-12	JCF15	22982	4.6311 -		SFAM01.1	PPB	9/16/2021 13:25	SMPL
CCV052	CCV052	23832	4.8041 -		SFAM01.1	PPB	9/16/2021 13:27	SMPL
CCB052	CCB052	162	-0.014 -		SFAM01.1	PPB	9/16/2021 13:29	SMPL