



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: jignesh

Date: 09/16/2021

OVEN IN TEMP1: 103 °C

OVEN OUT TEMP1: 104 °C

DATE & TIME IN1: 09/15/2021 16:00 DATE & TIME OUT1: 09/16/2021 08:00

OVEN IN TEMP2: 104 °C

OVEN OUT TEMP2: 104 °C

DATE & TIME IN2: 09/16/2021 08:30 DATE & TIME OUT2: 09/16/2021 10:00

Run Number: LB116345

BalanceID: Wc Sc-05

OvenID: Wc Oven-01

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB116345BL	LB116345BL	152.3891	152.3891	100	152.3892	152.3892	152.3892	0.0001	1
2	M3705-10	PT-SOL-WP	157.5161	157.5161	100	157.5895	157.5895	157.5895	0.0734	734

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$