

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : AL Instrument ID : Konelab

9/16/2021 14:07

Test: Total CN

log

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.356	0.0	0.071	
ICB1	1.817	0.0	0.001	
CCV1	239.953	0.0	0.182	
CCB1	1.603	0.0	0.001	
PB139047BL	1.777	0.0	0.001	
PB139047BS	96.633	0.0	0.073	
MIDPB139047	236.490	0.0	0.180	
M3705-23	330.260	0.0	0.251	
M3760-01	2.314	0.0	0.001	
M3760-01DUP	2.394	0.0	0.001	
M3760-01MS	36.540	0.0	0.027	
M3760-01MSD	36.658	0.0	0.028	
CCV2	242.568	0.0	0.184	
CCB2	1.510	0.0	0.001	

— ah y. (90-110)

N	14
Mean	94.634
SD	116.6785
CV%	123.29

Aquakem v. 7.2AQ1

Results from time period:

Thu Sep 16 13:54:10 2021

Thu Sep 16 14:05:09 2021

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.9354	µg/l	9/16/2021 8:44:03	
5.0PPBCN	A	Total CN	P	5.7854	µg/l	9/16/2021 8:44:04	
10PPBCN	A	Total CN	P	10.4836	µg/l	9/16/2021 8:44:05	
50PPBCN	A	Total CN	P	52.2694	µg/l	9/16/2021 8:44:06	
100PPBCN	A	Total CN	P	97.8746	µg/l	9/16/2021 8:44:07	
250PPBCN	A	Total CN	P	244.9418	µg/l	9/16/2021 8:44:08	
500PPBCN	A	Total CN	P	502.7097	µg/l	9/16/2021 8:44:09	
ICV1	S	Total CN	P	94.3565	µg/l	9/16/2021 13:54:11	
ICB1	S	Total CN	P	1.817	µg/l	9/16/2021 13:54:12	
CCV1	S	Total CN	P	239.953	µg/l	9/16/2021 13:54:14	
CCB1	S	Total CN	P	1.6026	µg/l	9/16/2021 13:54:17	
PB139047BL	S	Total CN	P	1.7773	µg/l	9/16/2021 13:54:18	
PB139047BS	S	Total CN	P	96.6327	µg/l	9/16/2021 13:54:20	
MIDPB139047	S	Total CN	P	236.4905	µg/l	9/16/2021 14:01:45	
M3705-23	S	Total CN	P	330.26	µg/l	9/16/2021 14:01:47	
M3760-01	S	Total CN	P	2.3145	µg/l	9/16/2021 14:01:48	
M3760-01DUP	S	Total CN	P	2.394	µg/l	9/16/2021 14:01:49	
M3760-01MS	S	Total CN	P	36.5401	µg/l	9/16/2021 14:01:50	
M3760-01MSD	S	Total CN	P	36.6582	µg/l	9/16/2021 14:01:51	
CCV2	S	Total CN	P	242.5678	µg/l	9/16/2021 14:01:54	
CCB2	S	Total CN	P	1.5104	µg/l	9/16/2021 14:05:08	

