

LB116350


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39602



A= 0.0000e+000

B= 2.5358e-004

C= -3.7998e-002

Rho= 0.9999238

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.012	-0.012	104	0.000	104					6
0.2	0.200	0.211	0.011	982	0.0 %	982					-2
2.5	2.500	2.462	-0.038	9858	0.0 %	9858					2
5.0	5.000	5.088	0.088	20216	0.0 %	20216					-1
7.5	7.500	7.446	-0.054	29514	0.0 %	29514					0
10.0	10.000	10.004	0.004	39602	0.0 %	39602					

LB116350 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Sep 2021 14:16:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	51	104	-	0.0000	7470A	PPB	16 Sep 2021 14:22:10	S	Std
0.2 - 1	51.2	982	-	0.2000	7470A	PPB	16 Sep 2021 14:24:27	S	Std
2.5 - 1	52.5	9858	-	2.5000	7470A	PPB	16 Sep 2021 14:26:43	S	Std
5.0 - 1	55	20216	-	5.0000	7470A	PPB	16 Sep 2021 14:29:00	S	Std
7.5 - 1	57.5	29514	-	7.5000	7470A	PPB	16 Sep 2021 14:31:16	S	Std
10.0 - 1	510	39602	-	10.0000	7470A	PPB	16 Sep 2021 14:33:32	S	Std
ICV94 - 1	ICV94	15123	3.7969	-	7470A	PPB	16 Sep 2021 14:36:18	U	SMPL
ICB94 - 1	ICB94	28	-0.0309	-	7470A	PPB	16 Sep 2021 14:38:32	U	SMPL
CCV53 - 1	CCV53	19738	4.9672	-	7470A	PPB	16 Sep 2021 14:40:47	U	SMPL
CCB53 - 1	CCB53	30	-0.0304	-	7470A	PPB	16 Sep 2021 14:43:02	U	SMPL
CRA - 1	CRA	983	0.2113	-	7470A	PPB	16 Sep 2021 14:45:17	U	SMPL
HighStd - 1	HighStd	38537	9.7342	-	7470A	PPB	16 Sep 2021 14:47:33	U	SMPL
ChkStd - 1	ChkStd	27529	6.9428	-	7470A	PPB	16 Sep 2021 14:49:48	U	SMPL
PB139149BL - 1	PBW	0	-0.0380	-	7470A	PPB	16 Sep 2021 14:52:05	U	SMPL
PB139149BS - 1	LCSW	14880	3.7353	-	7470A	PPB	16 Sep 2021 14:54:21	U	SMPL
M3742-01 - 1	TANK-1	186	0.0092	-	7470A	PPB	16 Sep 2021 14:56:37	U	SMPL
M3742-01DUP - 1	TANK-1DUP	203	0.0135	-	7470A	PPB	16 Sep 2021 14:58:53	U	SMPL
M3742-01MS - 1	TANK-1MS	13837	3.4708	-	7470A	PPB	16 Sep 2021 15:01:10	U	SMPL
M3742-01MSD - 1	TANK-1MSD	14287	3.5849	-	7470A	PPB	16 Sep 2021 15:03:27	U	SMPL
M3742-02 - 1	FRAC-TANK-NORTH	152	0.0005	-	7470A	PPB	16 Sep 2021 15:05:42	U	SMPL
CCV54 - 1	CCV54	20184	5.0803	-	7470A	PPB	16 Sep 2021 15:07:58	U	SMPL
CCB54 - 1	CCB54	30	-0.0304	-	7470A	PPB	16 Sep 2021 15:10:13	U	SMPL
CCV55 - 1	CCV55	19838	4.9925	-	7470A	PPB	17 Sep 2021 08:54:09	U	SMPL
CCB55 - 1	CCB55	106	-0.0111	-	7470A	PPB	17 Sep 2021 08:56:26	U	SMPL
PB139149TB - 1	PB139129TB	72	-0.0197	-	7470A	PPB	17 Sep 2021 08:58:42	U	SMPL
M3742-01LX5 - 1		0	-0.0380	-	7470A	PPB	17 Sep 2021 09:00:57	U	SMPL
M3742-01A - 1		15861	3.9840	-	7470A	PPB	17 Sep 2021 09:03:12	U	SMPL
CCV56 - 1		20249	5.0967	-	7470A	PPB	17 Sep 2021 09:05:28	U	SMPL
CCB56 - 1		9	-0.0357	-	7470A	PPB	17 Sep 2021 09:07:43	U	SMPL