



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116352

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	09/16/2021	09:45
2	M3728-01		1	10	100.00	>212.0	>212.0	09/16/2021	10:05
3	M3728-01DUP		1	10	100.00	>212.0	>212.0	09/16/2021	10:35
4	M3759-13		1	12	100.00	>212.0	>212.0	09/16/2021	11:10
5	M3759-14		1	10	100.00	>212.0	>212.0	09/16/2021	11:40
6	M3759-15		1	12	100.00	>212.0	>212.0	09/16/2021	12:05
7	M3759-17		1	14	100.00	>212.0	>212.0	09/16/2021	12:40
8	M3759-19		1	13	100.00	>212.0	>212.0	09/16/2021	13:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB116352

WORKLIST(Hardcopy Internal Chain)

WorkList Name: FP91621

WorkList ID: 152311

Department: Wet-Chemistry

Date: 09-16-2021 09:15:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/17/2021	Water	M3728-01	Flash Point	Cool 4 deg C	PSEG03	H51	N42049	09/14/2021	1010B
09/20/2021	Water	M3759-13	Flash Point	Cool 4 deg C	PSEG03	H51	1-2-3-5	09/15/2021	1010B
09/20/2021	Water	M3759-14	Flash Point	Cool 4 deg C	PSEG03	H51	6-7-8-9	09/15/2021	1010B
09/20/2021	Water	M3759-15	Flash Point	Cool 4 deg C	PSEG03	H51	41264	09/15/2021	1010B
09/20/2021	Water	M3759-17	Flash Point	Cool 4 deg C	PSEG03	H51	31185	09/15/2021	1010B
09/20/2021	Water	M3759-19	Flash Point	Cool 4 deg C	PSEG03	H51	41094	09/15/2021	1010B

Date/Time 09/16/21 09:25
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: JRC (WC)

Date/Time 09/16/21 13:15
Raw Sample Received by: JRC (WC)
Raw Sample Relinquished by: KS (WC)