



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116353

Slope : 98.7

BalanceID: Wc Sc-06

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	09/16/2021	15:15
2	CAL2	1	Water	NA	NA	20.6	7.00	09/16/2021	15:16
3	CAL3	1	Water	NA	NA	20.6	10.03	09/16/2021	15:17
4	ICV	1	Water	NA	NA	20.6	6.99	09/16/2021	15:19
5	CCV1	1	Water	NA	NA	20.7	2.01	09/16/2021	15:20
6	M3769-01	1	Solid	20.03	20	20.5	5.22	09/16/2021	16:00
7	M3773-01	1	Solid	20.04	20	22.1	7.65	09/16/2021	16:02
8	M3773-03	1	Solid	20.02	20	20.9	5.53	09/16/2021	16:05
9	M3773-03DUP	1	Solid	20.03	20	20.8	5.54	09/16/2021	16:06
10	CCV2	1	Water	NA	NA	20.7	12.03	09/16/2021	16:10

✓ 16353

WORKLIST(Hardcopy Internal Chain)

WorkList Name: PH M3769 S WorkList ID: 152319 Department: Wet-Chemistry Date: 09-16-2021 15:36:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/16/2021	Solid	M3769-01	pH	Cool 4 deg C	PSEG03	H61	NWB-1669	09/16/2021	9045D
09/16/2021	Solid	M3773-01	pH	Cool 4 deg C	PSEG03	H61	TRE-1203	09/16/2021	9045D
09/16/2021	Solid	M3773-03	pH	Cool 4 deg C	PSEG03	H61	TRE-1508	09/16/2021	9045D

Date/Time 09-16-21 15:40
Raw Sample Received by: JP (WCC) ITCsm
Raw Sample Relinquished by:

Date/Time 09-16-21 17:00
Raw Sample Received by: ITCsm
Raw Sample Relinquished by: JP (WCC)