



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116354

Slope : 98.7

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2810         |
| buffer solution pH 7 yellow    | W2813         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2809         |
| buffer solution pH 7 yellow    | W2719         |
| Buffer Solution, PH2 (500ml)   | W2806         |
| Buffer Solution, PH12 (500ml)  | W2807         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.6             | 4.01        | 09/16/2021 | 15:15     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.6             | 7.00        | 09/16/2021 | 15:16     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.6             | 10.03       | 09/16/2021 | 15:17     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.6             | 6.99        | 09/16/2021 | 15:19     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.7             | 2.01        | 09/16/2021 | 15:20     |
| 6   | M3769-03    | 1  | Water  | NA          | NA          | 23.1             | 7.14        | 09/16/2021 | 15:22     |
| 7   | M3776-01    | 1  | Water  | NA          | NA          | 21.6             | 5.10        | 09/16/2021 | 15:25     |
| 8   | M3776-01DUP | 1  | Water  | NA          | NA          | 21.7             | 5.11        | 09/16/2021 | 15:26     |
| 9   | CCV2        | 1  | Water  | NA          | NA          | 20.7             | 12.02       | 09/16/2021 | 15:30     |

18 116354

WORKLIST(Hardcopy Internal Chain)

WorkList Name: ph m3776      WorkList ID: 152316      Department: Wet-Chemistry      Date: 09-16-2021 15:05:01

| Due Date   | Matrix | Sample   | Test | Preservative | Customer | Raw Sample Storage Location | Customer Sample   | Collect Date | Method |
|------------|--------|----------|------|--------------|----------|-----------------------------|-------------------|--------------|--------|
| 09/16/2021 | Water  | M3769-03 | pH   | Cool 4 deg C | PSEG03   | H61                         | NWB-1670-1675     | 09/16/2021   | 9040C  |
| 09/10/2021 | Water  | M3776-01 | pH   | Cool 4 deg C | STER15   | D11                         | GILROY-OMEGA-CELL | 09/10/2021   | 9040C  |

Date/Time 09-16-21 15:10  
Raw Sample Received by: JP Lee  
Raw Sample Relinquished by: \_\_\_\_\_

Date/Time 09-16-21 14:00  
Raw Sample Received by: \_\_\_\_\_  
Raw Sample Relinquished by: JP Lee