

LB116364


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37840



A= 0.0000e+000

B= 2.6348e-004

C= -3.8407e-002

Rho= 0.9999206

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.009	-0.009	112	0.000	112					-
0.2	0.200	0.170	-0.030	791	0.0 %	791					15
2.5	2.500	2.490	-0.010	9596	0.0 %	9596					0
5.0	5.000	5.071	0.071	19393	0.0 %	19393					1
7.5	7.500	7.546	0.046	28787	0.0 %	28787					1
10.0	10.000	9.932	-0.068	37840	0.0 %	37840					-1

LB116364 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Sep 2021 11:11:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	112	-	0.0000	7470A	PPB	17 Sep 2021 11:12:47	S	Std
0.2 - 1	50.2	791	-	0.2000	7470A	PPB	17 Sep 2021 11:15:04	S	Std
2.5 - 1	52.5	9596	-	2.5000	7470A	PPB	17 Sep 2021 11:17:22	S	Std
5.0 - 1	55	19393	-	5.0000	7470A	PPB	17 Sep 2021 11:19:38	S	Std
7.5 - 1	57.5	28787	-	7.5000	7470A	PPB	17 Sep 2021 11:21:55	S	Std
10.0 - 1	510	37840	-	10.0000	7470A	PPB	17 Sep 2021 11:26:56	S	Std
ICV95 - 1	ICV95	15468	4.0370	-	7470A	PPB	17 Sep 2021 11:29:42	U	SMPL
ICB95 - 1	ICB95	-2	-0.0389	-	7470A	PPB	17 Sep 2021 11:31:57	U	SMPL
CCV57 - 1	CCV57	19010	4.9703	-	7470A	PPB	17 Sep 2021 11:34:12	U	SMPL
CCB57 - 1	CCB57	23	-0.0323	-	7470A	PPB	17 Sep 2021 11:36:27	U	SMPL
CRA - 1	CRA	784	0.1682	-	7470A	PPB	17 Sep 2021 11:38:42	U	SMPL
HighStd - 1	HighStd	37550	9.8551	-	7470A	PPB	17 Sep 2021 11:40:57	U	SMPL
ChkStd - 1	ChkStd	26235	6.8739	-	7470A	PPB	17 Sep 2021 11:43:12	U	SMPL
PB139171BL - 1	PBW	87	-0.0155	-	7470A	PPB	17 Sep 2021 11:48:00	U	SMPL
PB139171BS - 1	LCSW	14847	3.8734	-	7470A	PPB	17 Sep 2021 11:50:15	U	SMPL
M3774-01 - 1	SW-RR-02-(1.0)	-16	-0.0426	-	7470A	PPB	17 Sep 2021 11:52:31	U	SMPL
M3774-01DUP - 1	SW-RR-02-(1.0)DUP	22	-0.0326	-	7470A	PPB	17 Sep 2021 11:54:48	U	SMPL
M3774-01MS - 1	SW-RR-02-(1.0)MS	14739	3.8450	-	7470A	PPB	17 Sep 2021 11:57:05	U	SMPL
M3774-01MSD - 1	SW-RR-02-(1.0)MSD	14653	3.8223	-	7470A	PPB	17 Sep 2021 11:59:21	U	SMPL
M3774-02 - 1	SW-RR-02-(1.0)	-43	-0.0497	-	7470A	PPB	17 Sep 2021 12:01:37	U	SMPL
CCV58 - 1	CCV58	19060	4.9834	-	7470A	PPB	17 Sep 2021 12:03:52	U	SMPL
CCB58 - 1	CCB58	15	-0.0345	-	7470A	PPB	17 Sep 2021 12:06:07	U	SMPL
M3774-03 - 1	SW-CB-04-(1.5)	175	0.0077	-	7470A	PPB	17 Sep 2021 12:08:22	U	SMPL
M3774-04 - 1	SW-CB-04-(1.5)	16	-0.0342	-	7470A	PPB	17 Sep 2021 12:10:37	U	SMPL
M3774-05 - 1	SW-CB-02-(0.2)	11	-0.0355	-	7470A	PPB	17 Sep 2021 12:12:53	U	SMPL
M3774-06 - 1	SW-CB-02-(0.2)	-7	-0.0403	-	7470A	PPB	17 Sep 2021 12:15:09	U	SMPL
M3774-07 - 1	SW-RR-01-(2.0)	-6	-0.0400	-	7470A	PPB	17 Sep 2021 12:17:24	U	SMPL
M3774-08 - 1	SW-RR-01-(2.0)	63	-0.0218	-	7470A	PPB	17 Sep 2021 12:19:40	U	SMPL
M3774-09 - 1	SW-CB-03-(1.0)	21	-0.0329	-	7470A	PPB	17 Sep 2021 12:21:56	U	SMPL
M3774-10 - 1	SW-CB-03-(1.0)	3	-0.0376	-	7470A	PPB	17 Sep 2021 12:24:14	U	SMPL
M3774-01LX5 - 1		64	-0.0215	-	7470A	PPB	17 Sep 2021 12:26:30	U	SMPL
M3774-01A - 1		15582	4.0671	-	7470A	PPB	17 Sep 2021 12:28:46	U	SMPL
CCV59 - 1	CCV59	19122	4.9998	-	7470A	PPB	17 Sep 2021 12:31:02	U	SMPL
CCB59 - 1	CCB59	26	-0.0316	-	7470A	PPB	17 Sep 2021 12:33:18	U	SMPL