



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116365

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	09/16/2021	15:50
2	M3759-01		1	3	14.00	57.2	56.9	09/16/2021	16:15
3	M3759-01DUP		1	4	14.00	57.2	56.9	09/16/2021	16:32
4	M3759-03		1	8	60.00	140.0	139.7	09/16/2021	17:05
5	M3769-01		1	10	100.00	>212.0	>212.0	09/16/2021	17:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP91621-1 WorkList ID : 152312 Department : Wet-Chemistry Date : 09-16-2021 15:00:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/20/2021	Solid	M3759-01	Flash Point	Cool 4 deg C	PSEG03	H51	41095	09/15/2021	1010A
09/20/2021	Solid	M3759-03	Flash Point	Cool 4 deg C	PSEG03	H51	20859	09/15/2021	1010A
09/21/2021	Solid	M3769-01	Flash Point	Cool 4 deg C	PSEG03	H61	NWB-1669	09/16/2021	1010A

Date/Time 09/16/21 15:30
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JF (WC)

Date/Time 09/16/21 17:40
 Raw Sample Received by: JF (WC)
 Raw Sample Relinquished by: KS (WC)