

LB116367 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 17 Sep 2021 13:14:23

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	92	-	0.0000245.1	PPB	17 Sep 2021 13:16:16	S	Std
0.05 - 1	50.5	258	-	0.0500245.1	PPB	17 Sep 2021 13:18:33	S	Std
0.2 - 1	50.2	887	-	0.2000245.1	PPB	17 Sep 2021 13:20:51	S	Std
2.5 - 1	52.5	9133	-	2.5000245.1	PPB	17 Sep 2021 13:23:07	S	Std
5.0 - 1	55	18537	-	5.0000245.1	PPB	17 Sep 2021 13:25:24	S	Std
7.5 - 1	57.5	27375	-	7.5000245.1	PPB	17 Sep 2021 13:27:39	S	Std
10.0 - 1	510	37298	-	10.0000245.1	PPB	17 Sep 2021 13:29:55	S	Std
ICV96 - 1	ICV96	14362	3.8759	-245.1	PPB	17 Sep 2021 13:35:13	U	SMPL
ICB96 - 1	ICB96	30	-0.0018	-245.1	PPB	17 Sep 2021 13:37:28	U	SMPL
CCV60 - 1	CCV60	18757	5.0651	-245.1	PPB	17 Sep 2021 13:39:44	U	SMPL
CCB60 - 1	CCB60	36	-0.0002	-245.1	PPB	17 Sep 2021 13:41:58	U	SMPL
CRA - 1	CRA	260	0.0604	-245.1	PPB	17 Sep 2021 13:44:14	U	SMPL
HighStd - 1	HighStd	36820	9.9523	-245.1	PPB	17 Sep 2021 13:46:28	U	SMPL
ChkStd - 1	ChkStd	25621	6.9222	-245.1	PPB	17 Sep 2021 13:48:44	U	SMPL
PB139148BL - 1	PBW	14	-0.0062	-245.1	PPB	17 Sep 2021 13:51:00	U	SMPL
PB139148BS - 1	LCSW	13733	3.7057	-245.1	PPB	17 Sep 2021 13:55:50	U	SMPL
M3703-08 - 1	PT-HG-WP	20672	5.5832	-245.1	PPB	17 Sep 2021 13:58:06	U	SMPL
M3760-02 - 1	EFF-WW	18	-0.0051	-245.1	PPB	17 Sep 2021 14:00:23	U	SMPL
M3760-02DUP - 1	EFF-WWDUP	88	0.0139	-245.1	PPB	17 Sep 2021 14:02:39	U	SMPL
M3760-02MS - 1	EFF-WWMS	13697	3.6960	-245.1	PPB	17 Sep 2021 14:07:27	U	SMPL
M3760-02MSD - 1	EFF-WWMSD	13608	3.6719	-245.1	PPB	17 Sep 2021 14:09:42	U	SMPL
CCV61 - 1	CCV61	18504	4.9966	-245.1	PPB	17 Sep 2021 14:11:58	U	SMPL
CCB61 - 1	CCB61	41	0.0011	-245.1	PPB	17 Sep 2021 14:14:13	U	SMPL
M3760-02LX5 - 1		10	-0.0073	-245.1	PPB	17 Sep 2021 14:23:15	U	SMPL
M3760-02A - 1		13990	3.7753	-245.1	PPB	17 Sep 2021 14:25:31	U	SMPL
CCV62 - 1	CCV62	18820	5.0821	-245.1	PPB	17 Sep 2021 14:27:47	U	SMPL
CCB62 - 1	CCB62	35	-0.0005	-245.1	PPB	17 Sep 2021 14:30:03	U	SMPL