


LB116374

7470A

INSTRUMENT ID: CVL



Linear

A= 0.0000e+000

B= 2.5101e-004

C= -2.5094e-002

Rho= 0.9997885

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.020	-0.020	21	0.000	21					-
0.2	0.200	0.177	-0.023	807	0.0 %	807					-12
2.5	2.500	2.460	-0.040	9900	0.0 %	9900					-2
5.0	5.000	5.168	0.168	20689	0.0 %	20689					3
7.5	7.500	7.466	-0.034	29842	0.0 %	29842					0
10.0	10.000	9.949	-0.051	39735	0.0 %	39735					-1

LB116374 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Sep 2021 15:59:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	21	-	0.0000	7470A	PPB	17 Sep 2021 15:59:46	S	Std
0.2 - 1	50.2	807	-	0.2000	7470A	PPB	17 Sep 2021 16:02:03	S	Std
2.5 - 1	52.5	9900	-	2.5000	7470A	PPB	17 Sep 2021 16:04:20	S	Std
5.0 - 1	55	20689	-	5.0000	7470A	PPB	17 Sep 2021 16:06:37	S	Std
7.5 - 1	57.5	29842	-	7.5000	7470A	PPB	17 Sep 2021 16:08:53	S	Std
10.0 - 1	50	39735	-	10.0000	7470A	PPB	17 Sep 2021 16:11:09	S	Std
ICV98 - 1	ICV98	15208	3.7923	-	7470A	PPB	17 Sep 2021 16:13:48	U	SMPL
ICB98 - 1	ICB98	-25	-0.0314	-	7470A	PPB	17 Sep 2021 16:16:03	U	SMPL
CCV65 - 1	CCV65	20765	5.1871	-	7470A	PPB	17 Sep 2021 16:18:18	U	SMPL
CCB65 - 1	CCB65	-34	-0.0336	-	7470A	PPB	17 Sep 2021 16:20:33	U	SMPL
CRA - 1	CRA	783	0.1714	-	7470A	PPB	17 Sep 2021 16:22:48	U	SMPL
HighStd - 1	HighStd	39575	9.9087	-	7470A	PPB	17 Sep 2021 16:25:02	U	SMPL
ChkStd - 1	ChkStd	28451	7.1164	-	7470A	PPB	17 Sep 2021 16:27:17	U	SMPL
PB139182BL - 1	PBW	-48	-0.0371	-	7470A	PPB	17 Sep 2021 16:29:33	U	SMPL
PB139182BS - 1	LCSW	14616	3.6437	-	7470A	PPB	17 Sep 2021 16:31:49	U	SMPL
M3791-01 - 1	SW-RR-02-(1.0)	148	0.0121	-	7470A	PPB	17 Sep 2021 16:34:05	U	SMPL
M3791-01DUP - 1	SW-RR-02-(1.0)DUP	164	0.0161	-	7470A	PPB	17 Sep 2021 16:36:21	U	SMPL
M3791-01MS - 1	SW-RR-02-(1.0)MS	14138	3.5237	-	7470A	PPB	17 Sep 2021 16:38:39	U	SMPL
M3791-01MSD - 1	SW-RR-02-(1.0)MSD	14256	3.5533	-	7470A	PPB	17 Sep 2021 16:40:57	U	SMPL
M3791-02 - 1	SW-CB-03-(1.0)	33	-0.0168	-	7470A	PPB	17 Sep 2021 16:43:12	U	SMPL
CCV66 - 1	CCV66	20740	5.1809	-	7470A	PPB	17 Sep 2021 16:45:28	U	SMPL
CCB66 - 1	CCB66	-27	-0.0319	-	7470A	PPB	17 Sep 2021 16:47:43	U	SMPL
M3791-03 - 1	SW-CB-02-(0.3)	33	-0.0168	-	7470A	PPB	17 Sep 2021 16:49:58	U	SMPL
M3791-04 - 1	SW-RR-01-(2.0)	6	-0.0236	-	7470A	PPB	17 Sep 2021 16:52:13	U	SMPL
M3791-05 - 1	SW-CB-04-(1.5)	163	0.0158	-	7470A	PPB	17 Sep 2021 16:54:29	U	SMPL
M3791-01LX5 - 1		39	-0.0153	-	7470A	PPB	17 Sep 2021 16:56:45	U	SMPL
M3791-01A - 1		15373	3.8337	-	7470A	PPB	17 Sep 2021 16:59:01	U	SMPL
CCV67 - 1	CCV67	20836	5.2050	-	7470A	PPB	17 Sep 2021 17:01:17	U	SMPL
CCB67 - 1	CCB67	-15	-0.0289	-	7470A	PPB	17 Sep 2021 17:03:33	U	SMPL