



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116379

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	09/17/2021	14:35
2	CAL2	1	Water	NA	NA	20.6	7.00	09/17/2021	14:36
3	CAL3	1	Water	NA	NA	20.6	10.02	09/17/2021	14:37
4	ICV	1	Water	NA	NA	20.7	6.99	09/17/2021	14:39
5	CCV1	1	Water	NA	NA	20.6	2.01	09/17/2021	14:40
6	M3786-07	1	Water	NA	NA	21.5	7.74	09/17/2021	14:42
7	M3786-07DUP	1	Water	NA	NA	21.6	7.75	09/17/2021	14:43
8	CCV2	1	Water	NA	NA	20.8	12.02	09/17/2021	14:45

116379

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY M3786 WorkList ID : 152354 Department : Wet-Chemistry Date : 09-17-2021 14:20:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/17/2021	Water	M3786-07	Corrosivity	Cool 4 deg C	PSEG03	H62	N71924	09/17/2021	9040C

Date/Time 09-17-21 14:30
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (WC)

Date/Time 09-17-21 14:00
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (WC)