



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116380

Slope : 98.1

BalanceID: Wc Sc-06

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	09/17/2021	14:35
2	CAL2	1	Water	NA	NA	20.6	7.00	09/17/2021	14:36
3	CAL3	1	Water	NA	NA	20.6	10.02	09/17/2021	14:37
4	ICV	1	Water	NA	NA	20.7	6.99	09/17/2021	14:39
5	CCV1	1	Water	NA	NA	20.6	2.01	09/17/2021	14:40
6	M3762-02	1	Solid	20.03	20	22.5	7.76	09/17/2021	16:00
7	M3762-02DUP	1	Solid	20.04	20	22.6	7.77	09/17/2021	16:01
8	M3772-02	1	Solid	20.05	20	21.6	7.36	09/17/2021	16:10
9	M3772-04	1	Solid	20.04	20	21.4	7.18	09/17/2021	16:15
10	M3786-02	1	Solid	20.02	20	22.5	8.60	09/17/2021	16:20
11	M3786-04	1	Solid	20.03	20	22.2	7.88	09/17/2021	16:22
12	M3786-06	1	Solid	20.05	20	21.5	7.74	09/17/2021	16:25
13	CCV2	1	Water	NA	NA	20.8	12.02	09/17/2021	16:26

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M3772 WorkList ID : 152315 Department : Wet-Chemistry Date : 09-17-2021 15:03:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/15/2021	Solid	M3772-02	Corrosivity	Cool 4 deg C	ENV127	H61	PRE-CLASS-WASTE-1	09/15/2021	9045D
09/15/2021	Solid	M3772-04	Corrosivity	Cool 4 deg C	ENV127	H61	WASTE-PRE-CLASS	09/15/2021	9045D
09/16/2021	Solid	M3762-02	Corrosivity	Cool 4 deg C	PSEG03	H61	11934	09/16/2021	9045D
09/17/2021	Solid	M3786-02	Corrosivity	Cool 4 deg C	PSEG03	H62	CHRT-26878	09/17/2021	9045D
09/17/2021	Solid	M3786-04	Corrosivity	Cool 4 deg C	PSEG03	H62	CHRT-25949	09/17/2021	9045D
09/17/2021	Solid	M3786-06	Corrosivity	Cool 4 deg C	PSEG03	H62	CHRT-26414	09/17/2021	9045D

Date/Time 09-17-21 15:20
 Raw Sample Received by: JO (WLC)
 Raw Sample Relinquished by: JTCsm

Date/Time 09-17-21 17:30
 Raw Sample Received by: JTCsm
 Raw Sample Relinquished by: JO (WLC)