



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116381

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.4	4.01	09/17/2021	17:35
2	CAL2	1	Water	NA	NA	21.4	7.00	09/17/2021	17:36
3	CAL3	1	Water	NA	NA	21.1	10.03	09/17/2021	17:37
4	ICV	1	Water	NA	NA	21.1	6.99	09/17/2021	17:39
5	CCV1	1	Water	NA	NA	21.2	2.01	09/17/2021	17:40
6	M3792-05	1	Water	NA	NA	21.9	5.69	09/17/2021	17:42
7	M3792-05DUP	1	Water	NA	NA	21.8	5.70	09/17/2021	17:43
8	CCV2	1	Water	NA	NA	21.3	12.02	09/17/2021	17:45

2B 116381

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m3792 WorkList ID : 152366 Department : Wet-Chemistry Date : 09-17-2021 17:27:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/17/2021	Water	M3792-05	pH	Cool 4 deg C	PSEG03	H61	FM782	09/17/2021	9040C

Date/Time 09.17.21 17:20
Raw Sample Received by: JA WCC
Raw Sample Relinquished by: JA WCC

Date/Time 09.17.21 18:00
Raw Sample Received by: JA WCC
Raw Sample Relinquished by: JA WCC