



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116390

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	79.9	09/20/2021	09:35
2	M3769-03		1	10	100.00	>212.0	>212.0	09/20/2021	10:10
3	M3769-03DUP		1	10	100.00	>212.0	>212.0	09/20/2021	10:35
4	M3786-07		1	12	100.00	>212.0	>212.0	09/20/2021	11:05
5	M3792-05		1	11	100.00	>212.0	>212.0	09/20/2021	11:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB116390

WorkList Name : FP91721 WorkList ID : 152365 Department : Wet-Chemistry Date : 09-17-2021 16:26:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/30/2021	Water	M3769-03	Flash Point	Cool 4 deg C	PSEG03	H61	NWB-1670-1675	09/16/2021	1010B
09/22/2021	Water	M3786-07	Flash Point	Cool 4 deg C	PSEG03	H62	N71924	09/17/2021	1010B
09/24/2021	Water	M3792-05	Flash Point	Cool 4 deg C	PSEG03	H61	FM782	09/17/2021	1010B

Date/Time 09/20/21 09:00
 Raw Sample Received by: KS (wc)
 Raw Sample Relinquished by: CL sm

Date/Time 09/20/21 11:45
 Raw Sample Received by: CP sm
 Raw Sample Relinquished by: KS (wc)