

Analysis Method: 9034  
 Parameter: Reactive Sulfide  
 Run Number: LB116393

ANALYST: ketankumar  
 SUPERVISOR REVIEW BY: apatel  
 Constant: 16000  
 Normality1: 0.025  
 Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2843
IODINE SOLUTION .025N 1L	W2763
Starch Solution, 4L	W2703

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB139206BL		1	50	50	2.00	0.00	1.92	1.92	0.08	0.00	0.00	09/20/2021	12:10
2	M3786-07		1	50	50	2.00	1.92	3.80	1.88	0.12	0.04	0.32	09/20/2021	12:12
3	M3786-07DUP		1	50	50	2.00	3.80	5.70	1.90	0.10	0.02	0.16	09/20/2021	12:15

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 \* Normality1) - ((T1 - Value Corrected With Blank) \* Normality2)) \* Constant / Initial Volume