



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 9/21/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 09/20/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 09/21/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB116394

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M3816-01	MW-14-20210916-0-1	1	1.15	9.72	9.09	92.6	
M3816-02	MW-14-20210916-1-2-WC	2	1.18	9.6	8.98	92.6	
M3819-01	PCB-09321-SR-01	3	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3819-02	PCB-09321-SR-02	4	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3819-03	PCB-09321-SR-03	5	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3819-04	PCB-09321-SR-04	6	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3820-01	TB-CONCRETE	7	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M3821-01	TEST-PIT-9	8	1.18	9.77	8.34	83.4	
M3821-02	TEST-PIT-9-EPH-2	9	1.13	9.64	8.23	83.4	
M3823-01	DKM-ENV-IG-COMP	10	1.18	9.77	9.4	95.7	
M3823-03	DKM-ENV-IG-GRAB	11	1.16	9.97	9.48	94.4	
M3824-01	TP-B	12	1.13	9.67	8.81	89.9	
M3824-03	TP-B-EPH-2	13	1.12	9.63	8.85	90.8	
M3824-04	TP-C	14	1.15	9.95	9.16	91.0	
M3824-06	TP-C-EPH-2	15	1.16	9.7	8.97	91.5	
M3828-01	MW-22-20210920-0-1-WC	16	1.14	9.82	9.11	91.8	
M3828-02	MW-22-20210920-1-2-WC	17	1.16	9.86	8.5	84.4	
M3828-03	MW-22-20210920-2-3-WC	18	1.12	9.89	9.06	90.5	
M3828-04	MW-22-20210920-3-4-WC	19	1.1	9.88	8.78	87.5	
M3828-05	MW-22-20210920-4-5-WC	20	1.14	9.65	8.64	88.1	
M3828-05D	MW-22-20210920-4-5-WCDUP	21	1.16	9.67	8.66	88.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$