

LB116398


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38760



A= 0.0000e+000

B= 2.6091e-004

C= -8.4589e-004

Rho= 0.9994039

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	901
0.0	0.000	0.010	0.010	41	0.000	41					-7
0.2	0.200	0.187	-0.013	721	0.0 %	721					0
2.5	2.500	2.511	0.011	9627	0.0 %	9627					3
5.0	5.000	5.137	0.137	19691	0.0 %	19691					-3
7.5	7.500	7.243	-0.257	27766	0.0 %	27766					1
10.0	10.000	10.112	0.112	38760	0.0 %	38760					

LB116398 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 21 Sep 2021 09:01:05					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	41	-	0.0000	7470A	PPB	21 Sep 2021 09:11:45	S	Std
0.2 - 1	50.2	721	-	0.2000	7470A	PPB	21 Sep 2021 09:14:01	S	Std
2.5 - 1	52.5	9627	-	2.5000	7470A	PPB	21 Sep 2021 09:16:18	S	Std
5.0 - 1	55	19691	-	5.0000	7470A	PPB	21 Sep 2021 09:18:35	S	Std
7.5 - 1	57.5	27766	-	7.5000	7470A	PPB	21 Sep 2021 09:20:51	S	Std
10.0 - 1	510	38760	-	10.0000	7470A	PPB	21 Sep 2021 09:23:07	S	Std
ICV02 - 1	ICV02	14625	3.8149	-	7470A	PPB	21 Sep 2021 09:30:07	U	SMPL
ICB02 - 1	ICB02	-12	-0.0040	-	7470A	PPB	21 Sep 2021 09:32:22	U	SMPL
CCV78 - 1	CCV78	19503	5.0876	-	7470A	PPB	21 Sep 2021 09:34:37	U	SMPL
CCB78 - 1	CCB78	-6	-0.0024	-	7470A	PPB	21 Sep 2021 09:36:52	U	SMPL
CRA - 1	CRA	723	0.1878	-	7470A	PPB	21 Sep 2021 09:39:06	U	SMPL
HighStd - 1	HighStd	39451	10.2922	-	7470A	PPB	21 Sep 2021 09:41:21	U	SMPL
ChkStd - 1	ChkStd	26363	6.8774	-	7470A	PPB	21 Sep 2021 09:43:37	U	SMPL
PB139225BL - 1	PBW	-36	-0.0102	-	7470A	PPB	21 Sep 2021 09:45:52	U	SMPL
PB139225BS - 1	LCSW	14781	3.8556	-	7470A	PPB	21 Sep 2021 09:48:08	U	SMPL
M3730-05 - 1	WALL-BRICK-FLOOR	-2	-0.0014	-	7470A	PPB	21 Sep 2021 09:50:24	U	SMPL
M3730-05DUP - 1	WALL-BRICK-FLOORDUP	30	0.0070	-	7470A	PPB	21 Sep 2021 09:52:40	U	SMPL
M3730-05MS - 1	WALL-BRICK-FLOORMS	14754	3.8486	-	7470A	PPB	21 Sep 2021 09:54:57	U	SMPL
M3730-05MSD - 1	WALL-BRICK-FLOORMSD	16546	4.3161	-	7470A	PPB	21 Sep 2021 09:57:14	U	SMPL
M3762-02 - 1	11934	-13	-0.0042	-	7470A	PPB	21 Sep 2021 09:59:29	U	SMPL
CCV79 - 1	CCV79	19536	5.0962	-	7470A	PPB	21 Sep 2021 10:01:44	U	SMPL
CCB79 - 1	CCB79	-11	-0.0037	-	7470A	PPB	21 Sep 2021 10:03:59	U	SMPL
M3772-02 - 1	PRE-CLASS-WASTE-1	26	0.0059	-	7470A	PPB	21 Sep 2021 10:06:15	U	SMPL
M3772-04 - 1	WASTE-PRE-CLASS	39	0.0093	-	7470A	PPB	21 Sep 2021 10:08:31	U	SMPL
M3773-01 - 1	TRE-1203	29	0.0067	-	7470A	PPB	21 Sep 2021 10:10:47	U	SMPL
M3773-03 - 1	TRE-1508	22	0.0049	-	7470A	PPB	21 Sep 2021 10:13:02	U	SMPL
M3786-02 - 1	CHRT-26878	28	0.0065	-	7470A	PPB	21 Sep 2021 10:15:18	U	SMPL
M3786-04 - 1	CHRT-25949	31	0.0072	-	7470A	PPB	21 Sep 2021 10:17:34	U	SMPL
M3786-06 - 1	CHRT-26414	40	0.0096	-	7470A	PPB	21 Sep 2021 10:19:50	U	SMPL
M3803-03 - 1	212	61	0.0151	-	7470A	PPB	21 Sep 2021 10:22:07	U	SMPL
M3803-06 - 1	26160	132	0.0336	-	7470A	PPB	21 Sep 2021 10:24:23	U	SMPL
PB139226BL - 1	PBW	29	0.0067	-	7470A	PPB	21 Sep 2021 10:26:39	U	SMPL
CCV80 - 1	CCV80	19535	5.0960	-	7470A	PPB	21 Sep 2021 10:28:54	U	SMPL
CCB80 - 1	CCB80	-13	-0.0042	-	7470A	PPB	21 Sep 2021 10:31:10	U	SMPL
PB139226BS - 1	LCSW	15649	4.0821	-	7470A	PPB	21 Sep 2021 10:33:25	U	SMPL
M3769-03 - 1	NWB-1670-1675	-11	-0.0037	-	7470A	PPB	21 Sep 2021 10:35:41	U	SMPL
M3769-03DUP - 1	NWB-1670-1675DUP	22	0.0049	-	7470A	PPB	21 Sep 2021 10:37:56	U	SMPL
M3769-03MS - 1	NWB-1670-1675MS	15101	3.9391	-	7470A	PPB	21 Sep 2021 10:40:11	U	SMPL
M3769-03MSD - 1	NWB-1670-1675MSD	14477	3.7763	-	7470A	PPB	21 Sep 2021 10:42:27	U	SMPL
PB139227BL - 1	PBW	-7	-0.0027	-	7470A	PPB	21 Sep 2021 10:44:43	U	SMPL
PB139227BS - 1	LCSW	15209	3.9673	-	7470A	PPB	21 Sep 2021 10:46:59	U	SMPL
M3817-01 - 1	SW-RR-02-(1.0)	154	0.0393	-	7470A	PPB	21 Sep 2021 10:49:16	U	SMPL
M3817-01DUP - 1	SW-RR-02-(1.0)DUP	169	0.0432	-	7470A	PPB	21 Sep 2021 10:51:33	U	SMPL
M3817-01MS - 1	SW-RR-02-(1.0)MS	15780	4.1162	-	7470A	PPB	21 Sep 2021 10:53:48	U	SMPL
CCV81 - 1	CCV81	19928	5.1985	-	7470A	PPB	21 Sep 2021 10:56:04	U	SMPL
CCB81 - 1	CCB81	-28	-0.0082	-	7470A	PPB	21 Sep 2021 10:58:19	U	SMPL
M3817-01MSD - 1	SW-RR-02-(1.0)MSD	16290	4.2493	-	7470A	PPB	21 Sep 2021 11:00:35	U	SMPL
M3817-03 - 1	SW-CB-04-(1.5)	120	0.0305	-	7470A	PPB	21 Sep 2021 11:02:50	U	SMPL
M3817-05 - 1	SW-RR-01-(2.0)	12	0.0023	-	7470A	PPB	21 Sep 2021 11:05:06	U	SMPL
M3817-07 - 1	SW-CB-03-(1.0)	39	0.0093	-	7470A	PPB	21 Sep 2021 11:07:21	U	SMPL
M3817-09 - 1	SW-CB-02-(0.3)	26	0.0059	-	7470A	PPB	21 Sep 2021 11:09:37	U	SMPL
PB139225TB - 1	PB138152TB	32	0.0075	-	7470A	PPB	21 Sep 2021 11:11:53	U	SMPL
PB139226TB - 1	PB139188TB	36	0.0085	-	7470A	PPB	21 Sep 2021 11:14:09	U	SMPL
CCV82 - 1	CCV82	20132	5.2517	-	7470A	PPB	21 Sep 2021 11:16:27	U	SMPL
CCB82 - 1	CCB82	-13	-0.0042	-	7470A	PPB	21 Sep 2021 11:18:43	U	SMPL
M3730-05LX5 - 1		2	-0.0003	-	7470A	PPB	21 Sep 2021 11:20:59	U	SMPL
M3730-05A - 1		15807	4.1233	-	7470A	PPB	21 Sep 2021 11:23:15	U	SMPL
M3769-03LX5 - 1		-33	-0.0095	-	7470A	PPB	21 Sep 2021 11:25:30	U	SMPL
M3769-03A - 1		15806	4.1230	-	7470A	PPB	21 Sep 2021 11:27:45	U	SMPL
M3817-01LX5 - 1		-14	-0.0045	-	7470A	PPB	21 Sep 2021 11:30:01	U	SMPL
M3817-01A - 1		15795	4.1202	-	7470A	PPB	21 Sep 2021 11:32:17	U	SMPL
CCV83 - 1	CCV83	19372	5.0534	-	7470A	PPB	21 Sep 2021 11:34:32	U	SMPL
CCB83 - 1	CCB83	-24	-0.0071	-	7470A	PPB	21 Sep 2021 11:36:48	U	SMPL