

UB116399

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : AL

Instrument ID : Konelab

9/20/2021 14:06

Test: Phenolics- *mark*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.002	0.0	0.111	
ICB1	-0.006	0.0	0.006	
CCV1	0.997	0.0	0.110	
CCB1	-0.005	0.0	0.006	
PB139210BL	-0.006	0.0	0.006	
PB139210BS	1.001	0.0	0.110	
M3705-26	2.449	0.0	0.261	Test limit high
M3705-26DLX2	1.229	0.0	0.134	
CCV2	0.998	0.0	0.110	
CCB2	-0.002	0.0	0.006	

N	10
Mean	0.766
SD	0.7917
CV%	103.41

Aquakem v. 7.2AQ1

Results from time period:

Mon Sep 20 13:28:03 2021

Mon Sep 20 14:02:43 2021

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0021	mg/l	9/20/2021 13:28:03	
0.05PPM	A	Phenolics-	P	0.0537	mg/l	9/20/2021 13:28:04	
0.1PPM	A	Phenolics-	P	0.1055	mg/l	9/20/2021 13:28:05	
0.25PPM	A	Phenolics-	P	0.2514	mg/l	9/20/2021 13:28:06	
0.50PPM	A	Phenolics-	P	0.4998	mg/l	9/20/2021 13:28:07	
1.0PPM	A	Phenolics-	P	0.9758	mg/l	9/20/2021 13:28:08	
2.0PPM	A	Phenolics-	P	2.0116	mg/l	9/20/2021 13:28:09	
ICV1	S	Phenolics-	P	1.0019	mg/l	9/20/2021 13:55:50	
ICB1	S	Phenolics-	P	-0.006	mg/l	9/20/2021 13:55:52	
CCV1	S	Phenolics-	P	0.9971	mg/l	9/20/2021 13:55:55	
CCB1	S	Phenolics-	P	-0.0048	mg/l	9/20/2021 13:55:56	
PB139210BL	S	Phenolics-	P	-0.0065	mg/l	9/20/2021 13:55:59	
PB139210BS	S	Phenolics-	P	1.0015	mg/l	9/20/2021 13:56:01	
M3705-26	S	Phenolics-	P	2.4489	mg/l	9/20/2021 14:02:37	
M3705-26DLX2	S	Phenolics-	P	1.2289	mg/l	9/20/2021 14:02:39	
CCV2	S	Phenolics-	P	0.9979	mg/l	9/20/2021 14:02:41	
CCB2	S	Phenolics-	P	-0.0022	mg/l	9/20/2021 14:02:43	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : AS Instrument ID : Konelab

9/20/2021 13:30


Test Phenolics-

Accepted 9/20/2021 13:30

Factor 9.604
Bias 0.006

Coeff. of det. 0.999756

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.006	0.0021	0.0000	2.6
2	Phenol-2	0.012	0.0537	0.0500	5.5
3	Phenol-2	0.017	0.1055	0.1000	0.6
4	Phenol-2	0.032	0.2514	0.2500	0.0
5	Phenol-2	0.058	0.4998	0.5000	-2.4
6	Phenol-2	0.108	0.9758	1.0000	0.6
7	Phenol-2	0.216	2.0116	2.0000	

AL
9/20/21