



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116401

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/21/2021	10:20
2	CAL2	1	Water	NA	NA	20.7	7.00	09/21/2021	10:21
3	CAL3	1	Water	NA	NA	20.7	10.03	09/21/2021	10:22
4	ICV	1	Water	NA	NA	20.7	6.99	09/21/2021	10:24
5	CCV1	1	Water	NA	NA	20.6	2.01	09/21/2021	10:25
6	M3826-01	1	Water	NA	NA	20.1	5.75	09/21/2021	10:50
7	M3826-01DUP	1	Water	NA	NA	20.2	5.76	09/21/2021	10:51
8	M3826-03	1	Water	NA	NA	20.5	4.84	09/21/2021	10:55
9	M3826-05	1	Water	NA	NA	20.6	6.49	09/21/2021	10:59
10	M3826-07	1	Water	NA	NA	20.4	5.45	09/21/2021	11:00
11	M3826-09	1	Water	NA	NA	20.6	5.04	09/21/2021	11:05
12	M3826-10	1	Water	NA	NA	20.3	5.43	09/21/2021	11:11
13	M3826-11	1	Water	NA	NA	20.4	5.75	09/21/2021	11:15
14	M3826-12	1	Water	NA	NA	20.4	5.59	09/21/2021	11:20
15	CCV2	1	Water	NA	NA	20.5	12.02	09/21/2021	11:22

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3826 W

WorkList ID : 152427

Department : Wet-Chemistry

Date : 09-21-2021 10:15:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/20/2021	Water	M3826-01	pH	Cool 4 deg C	PSEG03	J11	AUD-1188-AUD-1189	09/20/2021	9040C
09/20/2021	Water	M3826-03	pH	Cool 4 deg C	PSEG03	J11	AUD-1183	09/20/2021	9040C
09/20/2021	Water	M3826-05	pH	Cool 4 deg C	PSEG03	J11	DRIP-COMP-AUD-1181-1	09/20/2021	9040C
09/20/2021	Water	M3826-07	pH	Cool 4 deg C	PSEG03	J11	AUD-1182	09/20/2021	9040C
09/20/2021	Water	M3826-09	pH	Cool 4 deg C	PSEG03	J11	1188	09/20/2021	9040C
09/20/2021	Water	M3826-10	pH	Cool 4 deg C	PSEG03	J11	1189	09/20/2021	9040C
09/20/2021	Water	M3826-11	pH	Cool 4 deg C	PSEG03	J11	1181	09/20/2021	9040C
09/20/2021	Water	M3826-12	pH	Cool 4 deg C	PSEG03	J11	1187	09/20/2021	9040C

Date/Time 09/21/21 10:15
 Raw Sample Received by: JR
 Raw Sample Relinquished by: CP sm

Date/Time 09/21/21 12:00
 Raw Sample Received by: CP sm
 Raw Sample Relinquished by: JR (WCC)