



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116415

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	09/21/2021	12:25
2	CAL2	1	Water	NA	NA	20.6	7.00	09/21/2021	12:26
3	CAL3	1	Water	NA	NA	20.6	10.03	09/21/2021	12:27
4	ICV	1	Water	NA	NA	20.7	7.01	09/21/2021	12:30
5	CCV1	1	Water	NA	NA	20.7	2.01	09/21/2021	12:32
6	M3836-01	1	Water	NA	NA	22.0	5.08	09/21/2021	12:35
7	M3836-01DUP	1	Water	NA	NA	22.1	5.09	09/21/2021	12:36
8	M3836-02	1	Water	NA	NA	22.2	5.21	09/21/2021	12:40
9	M3836-03	1	Water	NA	NA	22.3	5.28	09/21/2021	12:43
10	CCV2	1	Water	NA	NA	20.6	12.02	09/21/2021	12:45

WORKLIST(Hardcopy Internal Chain)

13116415

WorkList Name : tss m3636 WorkList ID : 152438 Department : Wet-Chemistry Date : 09-21-2021 12:18:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/16/2021	Water	M3836-01	pH	Cool 4 deg C	STER07	D11	CELL-A	09/16/2021	9040C
09/16/2021	Water	M3836-02	pH	Cool 4 deg C	STER07	D11	CELLS-B/C	09/16/2021	9040C
09/16/2021	Water	M3836-03	pH	Cool 4 deg C	STER07	D11	CELL-D	09/16/2021	9040C

Date/Time 09/21/21 12:20
 Raw Sample Received by: JR (w/c)
 Raw Sample Relinquished by: CP SM

Date/Time 09/21/21 13:35
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JR (w/c)