



Analytical Summary Report

Analysis Method: 1010B  
Parameter: Flash Point  
Run Number: LB116420  
Thermometer ID: flashpoint

Reviewed By: ketankumar  
Supervisor Review By: apatel  
Ambient Barometric Pressure (mmHg): 765.00  
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	09/21/2021	13:20
2	M3826-01		1	10	100.00	>212.0	>212.0	09/21/2021	13:45
3	M3826-01DUP		1	10	100.00	>212.0	>212.0	09/21/2021	14:10
4	M3826-03		1	12	100.00	>212.0	>212.0	09/21/2021	14:40
5	M3826-05		1	11	100.00	>212.0	>212.0	09/21/2021	15:05
6	M3826-07		1	10	100.00	>212.0	>212.0	09/21/2021	15:25

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$

$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$

WORKLIST(Hardcopy Internal Chain)

LB116420

WorkList Name : fp92121      WorkList ID : 152435      Department : Wet-Chemistry      Date : 09-21-2021 11:23:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/23/2021	Water	M3826-01	Flash Point	Cool 4 deg C	PSEG03	J11	AUD-1188-AUD-1189	09/20/2021	1010B
09/23/2021	Water	M3826-03	Flash Point	Cool 4 deg C	PSEG03	J11	AUD-1183	09/20/2021	1010B
09/23/2021	Water	M3826-05	Flash Point	Cool 4 deg C	PSEG03	J11	DRIP-COMP-AUD-1181-1	09/20/2021	1010B
09/23/2021	Water	M3826-07	Flash Point	Cool 4 deg C	PSEG03	J11	AUD-1182	09/20/2021	1010B

Date/Time 09/21/21 13:00  
 Raw Sample Received by: KS (WC)  
 Raw Sample Relinquished by: RJ (EXT-10th)

Date/Time 09/21/21 15:30  
 Raw Sample Received by: RJ (EXT-10th)  
 Raw Sample Relinquished by: KS (WC)