



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116424

Slope : 98.3

BalanceID: Ws Sc-04

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/21/2021	16:25
2	CAL2	1	Water	NA	NA	20.7	7.01	09/21/2021	16:26
3	CAL3	1	Water	NA	NA	20.7	10.03	09/21/2021	16:27
4	ICV	1	Water	NA	NA	20.8	6.99	09/21/2021	16:30
5	CCV1	1	Water	NA	NA	20.8	2.01	09/21/2021	16:32
6	M3832-02	1	Solid	20.03	20	22.8	9.23	09/21/2021	16:35
7	M3860-02	1	Solid	20.02	20	21.7	6.30	09/21/2021	16:36
8	M3860-02DUP	1	Solid	20.03	20	21.8	6.31	09/21/2021	16:37
9	CCV2	1	Water	NA	NA	20.7	12.02	09/21/2021	16:40

2B116424

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M3832 WorkList ID : 152447 Department : Wet-Chemistry Date : 09-21-2021 16:09:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/21/2021	Solid	M3832-02	Corrosivity	Cool 4 deg C	PSEG03	J12	TP-B	09/21/2021	9045D
09/21/2021	Solid	M3860-02	Corrosivity	Cool 4 deg C	PSEG03	J11	PT-SOIL	09/21/2021	9045D

Date/Time 09/21/21 16:10
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (WC)

Date/Time 09/21/21 17:00
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (WC)