


LB 116426

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.033	-0.033	-74	0.000	-74				
0.2	0.200	0.218	0.018	869	0.0 %	869				
2.5	2.500	2.367	-0.133	8946	0.0 %	8946				
5.0	5.000	5.177	0.177	19509	0.0 %	19509				
7.5	7.500	7.652	0.152	28810	0.0 %	28810				
10.0	10.000	9.819	-0.181	36954	0.0 %	36954				

4/6/11

-

9

-5

4

2

-2

LB116426 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 22 Sep 2021 07:55:38					
Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	-74	-	0.0000	7470A	PPB	22 Sep 2021 08:02:19	S	Std
0.2 - 1	50.2	869	-	0.2000	7470A	PPB	22 Sep 2021 08:04:36	S	Std
2.5 - 1	52.5	8946	-	2.5000	7470A	PPB	22 Sep 2021 08:06:53	S	Std
5.0 - 1	55	19509	-	5.0000	7470A	PPB	22 Sep 2021 08:11:44	S	Std
7.5 - 1	57.5	28810	-	7.5000	7470A	PPB	22 Sep 2021 08:14:01	S	Std
10.0 - 1	510	36954	-	10.0000	7470A	PPB	22 Sep 2021 08:21:25	S	Std
ICV04 - 1	ICV04	13967	3.7028	-	7470A	PPB	22 Sep 2021 08:24:15	U	SMPL
ICB04 - 1	ICB04	16	-0.0094	-	7470A	PPB	22 Sep 2021 08:26:30	U	SMPL
CCV89 - 1	CCV89	19602	5.2022	-	7470A	PPB	22 Sep 2021 08:28:45	U	SMPL
CCB89 - 1	CCB89	12	-0.0105	-	7470A	PPB	22 Sep 2021 08:30:59	U	SMPL
CRA - 1	CRA	834	0.2082	-	7470A	PPB	22 Sep 2021 08:33:14	U	SMPL
HighStd - 1	HighStd	36757	9.7669	-	7470A	PPB	22 Sep 2021 08:35:29	U	SMPL
ChkStd - 1	ChkStd	26699	7.0906	-	7470A	PPB	22 Sep 2021 08:37:44	U	SMPL
PB139250BL - 1	PBW	78	0.0071	-	7470A	PPB	22 Sep 2021 08:40:00	U	SMPL
PB139250BS - 1	LCSW	14309	3.7938	-	7470A	PPB	22 Sep 2021 08:42:16	U	SMPL
M3821-03 - 1	TEST-PIT-9	167	0.0307	-	7470A	PPB	22 Sep 2021 08:44:32	U	SMPL
M3823-02 - 1	DKM-ENV-IG-COMP	144	0.0246	-	7470A	PPB	22 Sep 2021 08:46:48	U	SMPL
M3823-04 - 1	DKM-ENV-IG-GRAB	137	0.0228	-	7470A	PPB	22 Sep 2021 08:49:05	U	SMPL
M3824-02 - 1	TP-B	152	0.0268	-	7470A	PPB	22 Sep 2021 08:51:23	U	SMPL
M3824-05 - 1	TP-C	481	0.1143	-	7470A	PPB	22 Sep 2021 08:53:38	U	SMPL
CCV90 - 1	CCV90	19465	5.1657	-	7470A	PPB	22 Sep 2021 08:55:53	U	SMPL
CCB90 - 1	CCB90	19	-0.0086	-	7470A	PPB	22 Sep 2021 08:58:09	U	SMPL
M3824-05DUP - 1	TP-CDUP	344	0.0778	-	7470A	PPB	22 Sep 2021 09:00:24	U	SMPL
M3824-05MS - 1	TP-CMS	16196	4.2959	-	7470A	PPB	22 Sep 2021 09:02:39	U	SMPL
M3824-05MSD - 1	TP-CMSD	14592	3.8691	-	7470A	PPB	22 Sep 2021 09:04:55	U	SMPL
PB139251BL - 1	PBW	4	-0.0126	-	7470A	PPB	22 Sep 2021 09:07:10	U	SMPL
PB139251BS - 1	LCSW	14822	3.9303	-	7470A	PPB	22 Sep 2021 09:09:26	U	SMPL
M3826-01 - 1	AUD-1188-AUD-1189	204	0.0406	-	7470A	PPB	22 Sep 2021 09:11:42	U	SMPL
M3826-03 - 1	AUD-1183	388	0.0895	-	7470A	PPB	22 Sep 2021 09:13:59	U	SMPL
M3826-05 - 1	DRIP-COMP-AUD-1181-1187	434	0.1018	-	7470A	PPB	22 Sep 2021 09:16:15	U	SMPL
M3826-07 - 1	AUD-1182	407	0.0946	-	7470A	PPB	22 Sep 2021 09:18:32	U	SMPL
M3826-07DUP - 1	AUD-1182DUP	384	0.0885	-	7470A	PPB	22 Sep 2021 09:20:47	U	SMPL
CCV91 - 1	CCV91	19354	5.1362	-	7470A	PPB	22 Sep 2021 09:23:03	U	SMPL
CCB91 - 1	CCB91	18	-0.0089	-	7470A	PPB	22 Sep 2021 09:25:18	U	SMPL
M3826-07MS - 1	AUD-1182MS	14495	3.8433	-	7470A	PPB	22 Sep 2021 09:30:20	U	SMPL
M3826-07MSD - 1	AUD-1182MSD	14737	3.9077	-	7470A	PPB	22 Sep 2021 09:32:35	U	SMPL
PB139252BL - 1	PBW	65	0.0036	-	7470A	PPB	22 Sep 2021 09:34:50	U	SMPL
PB139252BS - 1	LCSW	14168	3.7563	-	7470A	PPB	22 Sep 2021 09:37:06	U	SMPL
M3817-02 - 1	SW-RR-02-(1.0)	69	0.0047	-	7470A	PPB	22 Sep 2021 09:39:22	U	SMPL
M3817-04 - 1	SW-CB-04-(1.5)	274	0.0592	-	7470A	PPB	22 Sep 2021 09:41:38	U	SMPL
M3817-06 - 1	SW-RR-01-(2.0)	105	0.0142	-	7470A	PPB	22 Sep 2021 09:43:55	U	SMPL
M3817-08 - 1	SW-CB-03-(1.0)	108	0.0150	-	7470A	PPB	22 Sep 2021 09:46:12	U	SMPL
M3817-10 - 1	SW-CB-02-(0.3)	97	0.0121	-	7470A	PPB	22 Sep 2021 09:48:28	U	SMPL
M3835-01 - 1	SW-RR-02-(1.0)	816	0.2034	-	7470A	PPB	22 Sep 2021 09:50:44	U	SMPL
CCV92 - 1	CCV92	18322	4.8616	-	7470A	PPB	22 Sep 2021 09:53:00	U	SMPL
CCB92 - 1	CCB92	46	-0.0015	-	7470A	PPB	22 Sep 2021 09:55:15	U	SMPL
M3835-01DUP - 1	SW-RR-02-(1.0)DUP	830	0.2072	-	7470A	PPB	22 Sep 2021 10:00:18	U	SMPL
M3835-01MS - 1	SW-RR-02-(1.0)MS	15185	4.0269	-	7470A	PPB	22 Sep 2021 10:02:33	U	SMPL
M3835-01MSD - 1	SW-RR-02-(1.0)MSD	15468	4.1022	-	7470A	PPB	22 Sep 2021 10:04:49	U	SMPL
M3835-02 - 1	SW-CB-03-(1.0)	1051	0.2660	-	7470A	PPB	22 Sep 2021 10:07:05	U	SMPL
M3835-03 - 1	SW-CB-04-(1.0)	1724	0.4450	-	7470A	PPB	22 Sep 2021 10:09:20	U	SMPL
M3835-04 - 1	SW-CB-02-(0.2)	74	0.0060	-	7470A	PPB	22 Sep 2021 10:11:37	U	SMPL
M3835-05 - 1	SW-RR-01-(1.5)	188	0.0363	-	7470A	PPB	22 Sep 2021 10:13:53	U	SMPL
PB139250TB - 1	PB139229TB	70	0.0049	-	7470A	PPB	22 Sep 2021 10:16:10	U	SMPL
PB139251TB - 1	PB139244TB	180	0.0342	-	7470A	PPB	22 Sep 2021 10:18:27	U	SMPL
CCV93 - 1	CCV93	18522	4.9148	-	7470A	PPB	22 Sep 2021 10:20:42	U	SMPL
CCB93 - 1	CCB93	50	-0.0004	-	7470A	PPB	22 Sep 2021 10:22:58	U	SMPL
M3824-05LX5 - 1		81	0.0079	-	7470A	PPB	22 Sep 2021 10:25:13	U	SMPL
M3824-05A - 1		13619	3.6102	-	7470A	PPB	22 Sep 2021 10:27:29	U	SMPL
M3826-07LX5 - 1		24	-0.0073	-	7470A	PPB	22 Sep 2021 10:29:44	U	SMPL
M3826-07A - 1		13668	3.6232	-	7470A	PPB	22 Sep 2021 10:31:59	U	SMPL
M3835-01LX5 - 1		109	0.0153	-	7470A	PPB	22 Sep 2021 10:34:15	U	SMPL
M3835-01A - 1		16401	4.3504	-	7470A	PPB	22 Sep 2021 10:36:30	U	SMPL
CCV94 - 1	CCV94	19182	5.0904	-	7470A	PPB	22 Sep 2021 10:38:47	U	SMPL
CCB94 - 1	CCB94	24	-0.0073	-	7470A	PPB	22 Sep 2021 10:41:03	U	SMPL