

=====

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by :

Instrument ID : Konelab

9/22/2021 11:40

Test: *Detected*
~~Total~~ CN *unit*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.656	0.0	0.086	
ICB1	1.412	0.0	0.003	
CCV1	244.397	0.0	0.212	
CCB1	1.121	0.0	0.002	
M3772-02	1.132	0.0	0.002	
M3786-02	1.073	0.0	0.002	
M3786-06	1.164	0.0	0.002	
CCV2	248.718	0.0	0.216	
CCB2	0.230	0.0	0.002	

N	9
Mean	66.323
SD	106.9734
CV%	161.29

Aquakem v. 7.2AQ1

Results from time period:

Wed Sep 22 10:58:58 2021

Wed Sep 22 11:40:12 2021

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-0.2172	µg/l	9/22/2021 11:17:24	
5.0PPBCN	A	Reactive CN	P	4.2431	µg/l	9/22/2021 11:17:25	
10PPBCN	A	Reactive CN	P	9.1138	µg/l	9/22/2021 11:17:26	
50PPBCN	A	Reactive CN	P	49.0717	µg/l	9/22/2021 11:17:27	
100PPBCN	A	Reactive CN	P	100.9927	µg/l	9/22/2021 11:17:28	
250PPBCN	A	Reactive CN	P	253.7526	µg/l	9/22/2021 11:17:29	
500PPBCN	A	Reactive CN	P	498.0433	µg/l	9/22/2021 11:17:30	
ICV1	S	Reactive CN	P	97.6562	µg/l	9/22/2021 11:40:03	
ICB1	S	Reactive CN	P	1.4116	µg/l	9/22/2021 11:40:04	
CCV1	S	Reactive CN	P	244.3974	µg/l	9/22/2021 11:40:05	
CCB1	S	Reactive CN	P	1.1209	µg/l	9/22/2021 11:40:06	
M3772-02	S	Reactive CN	P	1.1317	µg/l	9/22/2021 11:40:07	
M3786-02	S	Reactive CN	P	1.0729	µg/l	9/22/2021 11:40:08	
M3786-06	S	Reactive CN	P	1.1643	µg/l	9/22/2021 11:40:09	
CCV2	S	Reactive CN	P	248.7183	µg/l	9/22/2021 11:40:10	
CCB2	S	Reactive CN	P	0.2305	µg/l	9/22/2021 11:40:11	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : AS

Instrument ID : Konelab

9/22/2021 11:18


Test Total CN

Accepted 9/22/2021 11:18

Factor 1159
Bias 0.001

Coeff. of det. 0.999897

Errors



	Calibrator	Response	Calc. con.	Conc.	Rel. Errors
1	0.0PPBCN	0.001	-0.2172	0.0000	-15.1
2	5.0PPBCN	0.005	4.2431	5.0000	-8.9
3	10PPBCN	0.009	9.1138	10.0000	-1.9
4	50PPBCN	0.044	49.0717	50.0000	1.0
5	100PPBCN	0.089	100.9927	100.0000	1.5
6	250PPBCN	0.220	253.7526	250.0000	-0.4
7	500PPBCN	0.431	498.0433	500.0000	

9/22/21