

LB116448


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

56674



A= 0.0000e+000

B= 1.7733e-004

C= 1.4796e-002

Rho= 0.9997455

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.049	0.049	195	0.000	195					16
0.2	0.200	0.231	0.031	1222	0.0 %	1222					0
2.5	2.500	2.512	0.012	14081	0.0 %	14081					-4
5.0	5.000	4.818	-0.182	27085	0.0 %	27085					0
7.5	7.500	7.525	0.025	42349	0.0 %	42349					1
10.0	10.000	10.065	0.065	56674	0.0 %	56674					

LB116448 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Sep 2021 15:47:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	195	-	0.0000	7471B	PPB	22 Sep 2021 15:50:59	S	Std
0.2 - 1	50.2	1222	-	0.2000	7471B	PPB	22 Sep 2021 15:55:23	S	Std
2.5 - 1	52.5	14081	-	2.5000	7471B	PPB	22 Sep 2021 15:57:39	S	Std
5.0 - 1	55	27085	-	5.0000	7471B	PPB	22 Sep 2021 15:59:56	S	Std
7.5 - 1	57.5	42349	-	7.5000	7471B	PPB	22 Sep 2021 16:02:13	S	Std
10.0 - 1	510	56674	-	10.0000	7471B	PPB	22 Sep 2021 16:04:29	S	Std
ICV07 - 1	ICV07	23257	4.1390	-	7471B	PPB	22 Sep 2021 16:07:18	U	SMPL
ICB07 - 1	ICB07	-45	0.0068	-	7471B	PPB	22 Sep 2021 16:09:33	U	SMPL
CCV01 - 1	CCV01	27170	4.8329	-	7471B	PPB	22 Sep 2021 16:11:48	U	SMPL
CCB01 - 1	CCB01	-57	0.0047	-	7471B	PPB	22 Sep 2021 16:14:03	U	SMPL
CRA - 1	CRA	1220	0.2311	-	7471B	PPB	22 Sep 2021 16:16:18	U	SMPL
HighStd - 1	HighStd	57172	10.1532	-	7471B	PPB	22 Sep 2021 16:18:33	U	SMPL
ChkStd - 1	ChkStd	39504	7.0201	-	7471B	PPB	22 Sep 2021 16:20:49	U	SMPL
PB139291BL - 1	PBS	-3	0.0143	-	7471B	PPB	22 Sep 2021 16:25:21	U	SMPL
PB139291BS - 1	LCSS	20490	3.6483	-	7471B	PPB	22 Sep 2021 16:27:37	U	SMPL
M3838-01 - 1	OR-02-092121	5974	1.0742	-	7471B	PPB	22 Sep 2021 16:29:53	U	SMPL
M3860-02 - 1	PT-SOIL	25166	4.4775	-	7471B	PPB	22 Sep 2021 16:32:10	U	SMPL
M3860-02DUP - 1	PT-SOILDUP	23463	4.1755	-	7471B	PPB	22 Sep 2021 16:34:27	U	SMPL
M3860-02MS - 1	PT-SOILMS	43934	7.8057	-	7471B	PPB	22 Sep 2021 16:36:44	U	SMPL
M3860-02MSD - 1	PT-SOILMSD	43445	7.7190	-	7471B	PPB	22 Sep 2021 16:38:59	U	SMPL
CCV02 - 1	CCV02	26978	4.7988	-	7471B	PPB	22 Sep 2021 16:41:15	U	SMPL
CCB02 - 1	CCB02	-56	0.0049	-	7471B	PPB	22 Sep 2021 16:43:31	U	SMPL
M3870-01 - 1	TP-C	5545	0.9981	-	7471B	PPB	22 Sep 2021 16:45:47	U	SMPL
M3871-01 - 1	SS01-BL-(0.0)	28318	5.0365	-	7471B	PPB	22 Sep 2021 16:48:02	U	SMPL
M3871-02 - 1	SS02-BL-(0.0)	26043	4.6330	-	7471B	PPB	22 Sep 2021 16:50:17	U	SMPL
M3871-03 - 1	SS03-BL-(0.0)	22912	4.0778	-	7471B	PPB	22 Sep 2021 16:52:33	U	SMPL
M3860-02LX5 - 1		5004	0.9022	-	7471B	PPB	22 Sep 2021 16:54:49	U	SMPL
M3860-02A - 1		46817	8.3169	-	7471B	PPB	22 Sep 2021 16:57:05	U	SMPL
CCV03 - 1	CCV03	26914	4.7875	-	7471B	PPB	22 Sep 2021 16:59:21	U	SMPL
CCB03 - 1	CCB03	-65	0.0033	-	7471B	PPB	22 Sep 2021 17:01:40	U	SMPL