



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116450

Slope : 98.7

BalanceID: Wc Sc-04

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/22/2021	15:30
2	CAL2	1	Water	NA	NA	20.7	7.00	09/22/2021	15:31
3	CAL3	1	Water	NA	NA	20.7	10.03	09/22/2021	15:33
4	ICV	1	Water	NA	NA	20.8	6.99	09/22/2021	15:35
5	CCV1	1	Water	NA	NA	20.7	2.01	09/22/2021	15:37
6	M3870-02	1	Solid	20.03	20	23.1	8.92	09/22/2021	15:40
7	M3870-02DUP	1	Solid	20.04	20	23.2	8.93	09/22/2021	15:41
8	M3872-02	1	Solid	20.02	20	23.5	8.87	09/22/2021	15:45
9	CCV2	1	Water	NA	NA	20.6	12.03	09/22/2021	15:47

16450

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY M3870 WorkList ID : 152491 Department : Wet-Chemistry Date : 09-22-2021 13:21:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/22/2021	Solid	M3870-02	Corrosivity	Cool 4 deg C	PSEG03	J21	TP-C	09/22/2021	9045D
09/22/2021	Solid	M3872-02	Corrosivity	Cool 4 deg C	PSEG03	J21	SP-LPA-01	09/22/2021	9045D

Date/Time 09/22/21 18:20
Raw Sample Received by: JR JWC
Raw Sample Relinquished by: JD CSM

Date/Time 09/22/21 17:00
Raw Sample Received by: JD CSM
Raw Sample Relinquished by: JR JWC