



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB116451

Analyst By : ketankumar

Supervisor Review By : apatel

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	09/22/2021	17:25
2	CAL2	1	Water	NA	NA	20.5	7.00	09/22/2021	17:26
3	CAL3	1	Water	NA	NA	20.6	10.02	09/22/2021	17:27
4	ICV	1	Water	NA	NA	20.7	6.99	09/22/2021	17:28
5	CCV1	1	Water	NA	NA	20.5	2.01	09/22/2021	17:29
6	M3868-02	1	Water	NA	NA	22.2	6.56	09/22/2021	17:30
7	M3868-02DUP	1	Water	NA	NA	22.3	6.55	09/22/2021	17:31
8	M3869-02	1	Water	NA	NA	22.3	5.50	09/22/2021	17:32
9	M3869-04	1	Water	NA	NA	22.3	8.60	09/22/2021	17:33
10	M3875-15	1	Water	NA	NA	24.1	7.25	09/22/2021	17:34
11	M3888-03	1	Water	NA	NA	24.9	4.58	09/22/2021	17:35
12	M3888-04	1	Water	NA	NA	24.4	4.56	09/22/2021	17:36
13	M3888-05	1	Water	NA	NA	24.8	5.15	09/22/2021	17:37
14	M3888-06	1	Water	NA	NA	24.8	5.95	09/22/2021	17:38
15	M3888-07	1	Water	NA	NA	25.3	6.45	09/22/2021	17:39
16	CCV2	1	Water	NA	NA	21.8	12.02	09/22/2021	17:40
17	M3888-08	1	Water	NA	NA	25.3	6.47	09/22/2021	17:41
18	M3888-09	1	Water	NA	NA	25.1	6.48	09/22/2021	17:42
19	M3888-10	1	Water	NA	NA	25.1	6.35	09/22/2021	17:43
20	M3888-11	1	Water	NA	NA	25.0	6.49	09/22/2021	17:44

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	M3888-12	1	Water	NA	NA	24.9	6.58	09/22/2021	17:45
22	M3888-13	1	Water	NA	NA	24.7	5.60	09/22/2021	17:46
23	CCV3	1	Water	NA	NA	21.6	2.01	09/22/2021	17:47

WORKLIST(Hardcopy Internal Chain)

LB 116457

WorkList Name : PH92221 WorkList ID : 152502 Department : Wet-Chemistry Date : 09-22-2021 17:20:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/22/2021	Water	M3868-02	pH	Cool 4 deg C	PSEG03	J21	33458	09/22/2021	9040C
09/22/2021	Water	M3869-02	pH	Cool 4 deg C	PSEG03	J21	33445	09/22/2021	9040C
09/22/2021	Water	M3869-04	pH	Cool 4 deg C	PSEG03	J21	33466	09/22/2021	9040C
09/22/2021	Water	M3875-15	pH	Cool 4 deg C	PSEG03	J21	20882	09/22/2021	9040C
09/22/2021	Water	M3888-03	pH	Cool 4 deg C	PSEG03	J22	20894	09/22/2021	9040C
09/22/2021	Water	M3888-04	pH	Cool 4 deg C	PSEG03	J22	20895	09/22/2021	9040C
09/22/2021	Water	M3888-05	pH	Cool 4 deg C	PSEG03	J22	20896	09/22/2021	9040C
09/22/2021	Water	M3888-06	pH	Cool 4 deg C	PSEG03	J22	20897	09/22/2021	9040C
09/22/2021	Water	M3888-07	pH	Cool 4 deg C	PSEG03	J22	40454	09/22/2021	9040C
09/22/2021	Water	M3888-08	pH	Cool 4 deg C	PSEG03	J22	40455	09/22/2021	9040C
09/22/2021	Water	M3888-09	pH	Cool 4 deg C	PSEG03	J22	40456	09/22/2021	9040C
09/22/2021	Water	M3888-10	pH	Cool 4 deg C	PSEG03	J22	40457	09/22/2021	9040C
09/22/2021	Water	M3888-11	pH	Cool 4 deg C	PSEG03	J22	40458	09/22/2021	9040C
09/22/2021	Water	M3888-12	pH	Cool 4 deg C	PSEG03	J22	40459	09/22/2021	9040C
09/22/2021	Water	M3888-13	pH	Cool 4 deg C	PSEG03	J22	20898	09/22/2021	9040C

Date/Time 09/22/21 17:22
 Raw Sample Received by: gs (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 09/22/21 18:20
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: gs (WC)