



### Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116452

Slope : 98.6

BalanceID:

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	09/22/2021	17:50
2	CAL2	1	Water	NA	NA	21.5	7.02	09/22/2021	17:51
3	CAL3	1	Water	NA	NA	20.7	10.02	09/22/2021	17:52
4	ICV	1	Water	NA	NA	20.6	7.01	09/22/2021	17:53
5	CCV1	1	Water	NA	NA	20.8	2.01	09/22/2021	17:54
6	M3875-03	1	Solid	20	20	23.1	5.07	09/22/2021	17:55
7	M3875-03DUP	1	Solid	20	20	23.1	5.08	09/22/2021	17:56
8	M3875-05	1	Solid	20	20	23.2	5.22	09/22/2021	17:57
9	M3875-07	1	Solid	20	20	23.3	5.29	09/22/2021	17:58
10	M3875-09	1	Solid	20	20	23.4	5.33	09/22/2021	17:59
11	M3875-11	1	Solid	20	20	23.7	7.29	09/22/2021	18:00
12	CCV2	1	Water	NA	NA	20.6	12.02	09/22/2021	18:01

WORKLIST(Hardcopy Internal Chain)

LB116452

WorkList Name : PH92221-1      WorkList ID : 152503      Department : Wet-Chemistry      Date : 09-22-2021 17:21:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/22/2021	Solid	M3875-03	pH	Cool 4 deg C	PSEG03	J21	20861	09/22/2021	9045D
09/22/2021	Solid	M3875-05	pH	Cool 4 deg C	PSEG03	J21	20862	09/22/2021	9045D
09/22/2021	Solid	M3875-07	pH	Cool 4 deg C	PSEG03	J21	20863	09/22/2021	9045D
09/22/2021	Solid	M3875-09	pH	Cool 4 deg C	PSEG03	J21	20860	09/22/2021	9045D
09/22/2021	Solid	M3875-11	pH	Cool 4 deg C	PSEG03	J21	20883	09/22/2021	9045D

Date/Time 09/22/21 17:22  
 Raw Sample Received by: JS (WC)  
 Raw Sample Relinquished by: JS (WC)

Date/Time 09/22/21  
 Raw Sample Received by: JS (WC)  
 Raw Sample Relinquished by: JS (WC)