



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 9/24/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 13:15
In Date: 09/23/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:48
Out Date: 09/24/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB116463

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M3845-01	EWY32	1	1.13	9.57	8.22	84.0	
M3845-02	EWY33	2	1.16	9.72	8.77	88.9	
M3845-03	EWY34	3	1.14	9.97	9.16	90.8	
M3845-04	EWY35	4	1.17	9.6	8.28	84.3	
M3845-05	EWY36	5	1.14	9.5	8.27	85.3	
M3845-06	EWY37	6	1.16	9.56	8.52	87.6	
M3845-07	EWY38	7	1.16	9.51	8.28	85.3	
M3845-08	EWY39	8	1.14	9.77	8.71	87.7	
M3845-09	EWY40	9	1.15	9.95	8.82	87.2	
M3845-10	EWY41	10	1.14	9.84	8.63	86.1	
M3845-11	EWY42	11	1.18	9.96	8.57	84.2	
M3845-12	EWY43	12	1.15	9.57	8.58	88.2	
M3845-13	EWY44	13	1.12	9.88	8.81	87.8	
M3845-14	EWY45	14	1.17	9.87	8.7	86.6	
M3845-15	EWY46	15	1.15	9.97	8.89	87.8	
M3845-16	EWY47	16	1.14	9.96	8.99	89.0	
M3845-17	EWY48	17	1.18	9.72	8.5	85.7	
M3845-18	EWY49	18	1.14	9.92	9.17	91.5	
M3845-19	EWY50	19	1.14	9.94	9.23	91.9	
M3845-20	EWY50D	20	1.14	9.94	9.23	91.9	
M3845-21	EWY50S	21	1.14	9.94	9.23	91.9	
M3845-22	EWY51	22	1.19	9.81	8.84	88.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$