



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116468

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.9	09/23/2021	09:40
2	M3875-03		1	8	56.00	132.8	133.1	09/23/2021	10:05
3	M3875-03DUP		1	8	55.00	131.0	131.3	09/23/2021	10:32
4	M3875-05		1	9	58.00	136.4	136.7	09/23/2021	10:58
5	M3875-07		1	9	57.00	134.6	134.9	09/23/2021	11:20
6	M3875-09		1	7	40.00	104.0	104.3	09/23/2021	11:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB116468

WorkList Name : fp92321 WorkList ID : 152526 Department : Wet-Chemistry Date : 09-23-2021 09:11:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/27/2021	Solid	M3875-03	Flash Point	Cool 4 deg C	PSEG03	J21	20861	09/22/2021	1010A
09/27/2021	Solid	M3875-05	Flash Point	Cool 4 deg C	PSEG03	J21	20862	09/22/2021	1010A
09/27/2021	Solid	M3875-07	Flash Point	Cool 4 deg C	PSEG03	J21	20863	09/22/2021	1010A
09/27/2021	Solid	M3875-09	Flash Point	Cool 4 deg C	PSEG03	J21	20860	09/22/2021	1010A

Date/Time 09/23/21 09:15
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 09/23/21 11:50
 Raw Sample Received by: JB (WC)
 Raw Sample Relinquished by: KS (WC)