



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116471

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.9	09/23/2021	09:40
2	M3869-02		1	10	100.00	>212.0	>212.0	09/23/2021	12:10
3	M3869-02DUP		1	10	100.00	>212.0	>212.0	09/23/2021	12:35
4	M3869-04		1	12	100.00	>212.0	>212.0	09/23/2021	13:05
5	M3875-15		1	11	100.00	>212.0	>212.0	09/23/2021	13:32
6	M3888-13		1	13	100.00	>212.0	>212.0	09/23/2021	13:58
7	M3888-14		1	12	100.00	>212.0	>212.0	09/23/2021	14:25
8	M3888-15		1	10	100.00	>212.0	>212.0	09/23/2021	14:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB116471

WorkList Name : fp92321-1 WorkList ID : 152527 Department : Wet-Chemistry Date : 09-23-2021 09:12:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/29/2021	Water	M3869-02	B Flash Point	Cool 4 deg C	PSEG03	J21	33445	09/22/2021	1010B
09/29/2021	Water	M3869-04	B Flash Point	Cool 4 deg C	PSEG03	J21	33466	09/22/2021	1010B
09/27/2021	Water	M3875-15	G Flash Point	Cool 4 deg C	PSEG03	J21	20882	09/22/2021	1010B
09/27/2021	Water	M3888-13	E Flash Point	Cool 4 deg C	PSEG03	J22	20898	09/22/2021	1010B
09/27/2021	Water	M3888-14	B Flash Point	Cool 4 deg C	PSEG03	J22	94-98	09/22/2021	1010B
09/27/2021	Water	M3888-15	B Flash Point	Cool 4 deg C	PSEG03	J22	54-59	09/22/2021	1010B

Date/Time 09/23/21 09:15
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJ (Ext-lab)

Date/Time 09/23/21 14:50
 Raw Sample Received by: RJ (Ext-lab)
 Raw Sample Relinquished by: KS (WC)