


7471B

LB116482  
INSTRUMENT ID: CV2

Linear

A= 0.0000e+000

B= 1.8827e-004

C= 2.9176e-002

Rho= 0.9999176

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | 0/0) |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|------|
| 0.0    | 0.000  | 0.058  | 0.058  | 153   | 0.000      | 153   |       |       |       |       |      |
| 0.2    | 0.200  | 0.182  | -0.018 | 810   | 0.0 %      | 810   |       |       |       |       | -9   |
| 2.5    | 2.500  | 2.460  | -0.040 | 12913 | 0.0 %      | 12913 |       |       |       |       | -2   |
| 5.0    | 5.000  | 5.018  | 0.018  | 26497 | 0.0 %      | 26497 |       |       |       |       | 0    |
| 7.5    | 7.500  | 7.429  | -0.071 | 39305 | 0.0 %      | 39305 |       |       |       |       | -1   |
| 10.0   | 10.000 | 10.053 | 0.053  | 53244 | 0.0 %      | 53244 |       |       |       |       | 1    |

## LB116482 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 23 Sep 2021 16:36:46

| Sample ID       | Extended ID  | μ Abs. | Conc.   | Std Conc     | Method | Units | Date                 | Type | Type |
|-----------------|--------------|--------|---------|--------------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 50           | 153    | -       | 0.00007471B  |        | PPB   | 23 Sep 2021 16:38:42 | S    | Std  |
| 0.2 - 1         | 50.2         | 810    | -       | 0.20007471B  |        | PPB   | 23 Sep 2021 16:43:41 | S    | Std  |
| 2.5 - 1         | 52.5         | 12913  | -       | 2.50007471B  |        | PPB   | 23 Sep 2021 16:45:58 | S    | Std  |
| 5.0 - 1         | 55           | 26497  | -       | 5.00007471B  |        | PPB   | 23 Sep 2021 16:48:14 | S    | Std  |
| 7.5 - 1         | 57.5         | 39305  | -       | 7.50007471B  |        | PPB   | 23 Sep 2021 16:50:31 | S    | Std  |
| 10.0 - 1        | 510          | 53244  | -       | 10.00007471B |        | PPB   | 23 Sep 2021 16:52:48 | S    | Std  |
| ICV10 - 1       | ICV10        | 21082  | 3.9982  | -7471B       |        | PPB   | 23 Sep 2021 16:56:26 | U    | SMPL |
| ICB10 - 1       | ICB10        | -39    | 0.0218  | -7471B       |        | PPB   | 23 Sep 2021 16:58:41 | U    | SMPL |
| CCV11 - 1       | CCV11        | 26902  | 5.0940  | -7471B       |        | PPB   | 23 Sep 2021 17:00:58 | U    | SMPL |
| CCB11 - 1       | CCB11        | -75    | 0.0151  | -7471B       |        | PPB   | 23 Sep 2021 17:03:13 | U    | SMPL |
| CRA - 1         | CRA          | 814    | 0.1824  | -7471B       |        | PPB   | 23 Sep 2021 17:05:29 | U    | SMPL |
| HighStd - 1     | HighStd      | 52967  | 10.0012 | -7471B       |        | PPB   | 23 Sep 2021 17:07:44 | U    | SMPL |
| ChkStd - 1      | ChkStd       | 37538  | 7.0964  | -7471B       |        | PPB   | 23 Sep 2021 17:09:59 | U    | SMPL |
| PB139313BL - 1  | PBS          | 104    | 0.0488  | -7471B       |        | PPB   | 23 Sep 2021 17:14:54 | U    | SMPL |
| PB139313BS - 1  | LCSS         | 21886  | 4.1496  | -7471B       |        | PPB   | 23 Sep 2021 17:17:09 | U    | SMPL |
| M3872-01 - 1    | SP-LPA-01    | 22789  | 4.3196  | -7471B       |        | PPB   | 23 Sep 2021 17:19:25 | U    | SMPL |
| M3875-11 - 1    | 20883        | 5148   | 0.9984  | -7471B       |        | PPB   | 23 Sep 2021 17:20:42 | U    | SMPL |
| M3875-13 - 1    | 40441        | 1152   | 0.2461  | -7471B       |        | PPB   | 23 Sep 2021 17:23:00 | U    | SMPL |
| M3881-01 - 1    | OK-01-092221 | 3580   | 0.7032  | -7471B       |        | PPB   | 23 Sep 2021 17:25:17 | U    | SMPL |
| M3886-01 - 1    | COMP-1       | 9851   | 1.8838  | -7471B       |        | PPB   | 23 Sep 2021 17:27:32 | U    | SMPL |
| CCV12 - 1       | CCV12        | 27599  | 5.2252  | -7471B       |        | PPB   | 23 Sep 2021 17:29:47 | U    | SMPL |
| CCB12 - 1       | CCB12        | -41    | 0.0215  | -7471B       |        | PPB   | 23 Sep 2021 17:32:02 | U    | SMPL |
| M3886-01DUP - 1 | COMP-1DUP    | 10939  | 2.0886  | -7471B       |        | PPB   | 23 Sep 2021 17:34:18 | U    | SMPL |
| M3886-01MS - 1  | COMP-1MS     | 31115  | 5.8871  | -7471B       |        | PPB   | 23 Sep 2021 17:39:22 | U    | SMPL |
| M3886-01MSD - 1 | COMP-1MSD    | 31865  | 6.0283  | -7471B       |        | PPB   | 23 Sep 2021 17:41:37 | U    | SMPL |
| M3886-02 - 1    | COMP-2       | 12879  | 2.4539  | -7471B       |        | PPB   | 23 Sep 2021 17:43:52 | U    | SMPL |
| M3886-01LX5 - 1 |              | 1919   | 0.3905  | -7471B       |        | PPB   | 23 Sep 2021 17:46:08 | U    | SMPL |
| M3886-01A - 1   |              | 31705  | 5.9982  | -7471B       |        | PPB   | 23 Sep 2021 17:48:24 | U    | SMPL |
| CCV13 - 1       | CCV13        | 27937  | 5.2888  | -7471B       |        | PPB   | 23 Sep 2021 17:50:40 | U    | SMPL |
| CCB13 - 1       | CCB13        | -37    | 0.0222  | -7471B       |        | PPB   | 23 Sep 2021 17:52:57 | U    | SMPL |