



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116485

Slope : 98.6

BalanceID:

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	09/23/2021	15:45
2	CAL2	1	Water	NA	NA	20.5	7.00	09/23/2021	15:46
3	CAL3	1	Water	NA	NA	20.5	10.02	09/23/2021	15:47
4	ICV	1	Water	NA	NA	20.6	6.99	09/23/2021	15:48
5	CCV1	1	Water	NA	NA	20.5	2.01	09/23/2021	15:49
6	M3868-05	1	Solid	20	20	22.2	6.56	09/23/2021	15:50
7	M3903-01	1	Solid	20	20	24.0	7.58	09/23/2021	15:51
8	M3903-03	1	Solid	20	20	24.1	9.08	09/23/2021	15:52
9	M3903-05	1	Solid	20	20	24.1	6.08	09/23/2021	15:53
10	M3903-07	1	Solid	20	20	24.2	6.25	09/23/2021	15:54
11	M3904-01	1	Solid	20	20	23.2	12.34	09/23/2021	15:55
12	M3904-01DUP	1	Solid	20	20	23.2	12.33	09/23/2021	15:56
13	M3906-01	1	Solid	20	20	23.4	8.66	09/23/2021	15:57
14	M3910-01	1	Solid	20	20	23.5	7.30	09/23/2021	15:58
15	M3910-03	1	Solid	20	20	23.8	7.45	09/23/2021	15:59
16	CCV2	1	Water	NA	NA	20.6	12.01	09/23/2021	16:00
17	M3914-01	1	Solid	20	20	23.9	9.35	09/23/2021	16:01
18	M3914-03	1	Solid	20	20	23.5	7.83	09/23/2021	16:02
19	CCV3	1	Water	NA	NA	20.5	2.01	09/23/2021	16:03

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m3904 WorkList ID : 152541 Department : Wet-Chemistry Date : 09-23-2021 14:02:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/22/2021	Solid	M3868-05	pH	Cool 4 deg C	PSEG03	J21	33458	09/22/2021	9045D
09/23/2021	Solid	M3903-01	pH	Cool 4 deg C	PSEG03	J22	225	09/23/2021	9045D
09/23/2021	Solid	M3903-03	pH	Cool 4 deg C	PSEG03	J22	RBR-254	09/23/2021	9045D
09/23/2021	Solid	M3903-05	pH	Cool 4 deg C	PSEG03	J22	11976	09/23/2021	9045D
09/23/2021	Solid	M3903-07	pH	Cool 4 deg C	PSEG03	J22	202	09/23/2021	9045D
09/23/2021	Solid	M3904-01	pH	Cool 4 deg C	TULL02	J23	JET-GROUT-SOIL-STEE	09/23/2021	9045D
09/23/2021	Solid	M3906-01	pH	Cool 4 deg C	PSEG03	J23	BS-DRUM-COMP	09/23/2021	9045D
09/23/2021	Solid	M3910-01	pH	Cool 4 deg C	PSEG03	J23	11986	09/23/2021	9045D
09/23/2021	Solid	M3910-03	pH	Cool 4 deg C	PSEG03	J23	208	09/23/2021	9045D
09/23/2021	Solid	M3914-01	pH	Cool 4 deg C	PSEG03	J23	LV-DRUM-SOIL	09/23/2021	9045D
09/23/2021	Solid	M3914-03	pH	Cool 4 deg C	PSEG03	J23	LV-DRUM-PADS	09/23/2021	9045D

Date/Time 09/23/21 15:40
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: KS (WC)

Date/Time 09/23/21 16:05
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: JP (WC)